Vivo 3, User's

ENGLISH (EU)

Vivo 3 User Manual

Breas Vivo 3
User Manual

007231 rev. 3.1 en-EU



Breas Medical AB, Företagsvägen 1, SE-435 33 Mölnlycke, Sweden
Phone +46 31 86 88 00
Order +46 31 86 88 20 , Technical Support +46 31 86 88 60
Fax +46 31 86 88 10, www.breas.com, Date: 2026-09-21 | 39923

CE 2797 UKCA 0086

7 321820 072316

Table of Contents

1 Introduction 5
1.1 Manufacturer Information 5
1.2 What is the Vivo 3? 6
1.2.1 Non Invasive Interfaces 6
1.2.2 Mobility and Usage Environment 6
1.2.3 Continuous Operation 6
1.2.4 Service Life 6
1.2.5 Cybersecurity 6
1.3 Intended Use 7
1.4 Clinicial Benefits 7
1.5 Operation by Lay Users 7
1.6 Contraindications 7
1.7 About this Manual 8
1.7.1 Audience 8
1.7.2 Icons in this Manual 8
2 Safety Information 10
2.1 General Use — Warnings and Precautions 10
2.2 Electricity — Warnings and Precautions 12
2.2.1 Electromagnetic Compatibility and Electrostatic Discharge (EMC and ESD) 13
2.3 Environment — Warnings and Precautions 13
2.4 Patient Circuit — Warnings and Precautions 14
2.5 Filter Usage — Warnings and Precautions 15
2.6 Invasive Use — Warnings and Precautions 16
2.7 Humidification and Heating — Warnings and Precautions 17
2.8 Cleaning and Maintenance — Warning and Precautions 18
2.9 Oxygen Usage — Warning and Precautions 19
2.10 Mobile Use — Warning and Precautions 20
3 Product Description 22
3.1 Overview 22
3.2 Ventilator Front 22
3.3 Ventilator Back 23
3.4 Power Management 23
3.5 Menus 24
3.5.1 Use the Menu 24
3.5.2 The Monitor Page 27
3.5.3 The Setup Pages 28
3.5.4 The Alarm Pages 28
3.5.5 The Others Pages 28
3.6 Symbols on the Vivo 3 31
3.7 Symbols on the Display 34

4 Treatment Functions and Settings 35
4.1 Treatment Settings 35
4.1.1 Fall Time 35
4.1.2 Ramp Up 35
4.1.3 Ramp Down 36
4.1.4 Humidifier 36
4.1.5 Humidifier Level 36
4.1.6 Heated Circuit 36
4.1.7 Heated Circuit Level 36
5 Prepare the Vivo 3 for Use 37
5.1 Checking the Vivo 3 before First Use 37
5.2 Placing the Vivo 3 37
5.3 Connecting the Vivo 3 to Power Supply 37
5.4 Connecting the Patient Circuit 38
5.4.1 Connect the Heated Patient Circuit 38
5.4.2 Disconnect the Heated Patient Circuit from the Ventilator 39
5.5 Performing Start-up Checks 39
5.6 Performing a Pre-Use Test 39
5.6.1 Actions At Pre-Use Test Failure 40
6 How to Use the Vivo 3 41
6.1 Switch On the Vivo 3 41
6.2 Start the Treatment 41
6.3 Stop the Treatment 42
6.4 Switch Off the Vivo 3 43
6.5 Monitor Treatment 44
6.5.1 Treatment Values Monitored by the Vivo 3 44
6.6 Using Accessories 44
6.6.1 Using the Attachable Humidifier 44
6.6.2 Using the Heated Circuit 47
6.6.3 Using the SpO_2 Module 48
6.6.4 Using the Oxygen Supply Adapter 49
6.6.5 Using the Nurse Call Connection 50
6.6.6 Using the Remote Alarm Unit 50
6.6.7 Using the Protective Cover 50
6.6.8 Using the Lightweight Mobility Bag 51
6.6.9 Using the Y-Cable 51
6.6.10 Using the Vivo 3 with the Trolley 51
6.7 Basic Troubleshooting 52
7 Alarms 53
7.1 Operator's Position 53
7.1.1 Checking the Operator's Position 54

7.2 Handle Alarms 54
7.2.1 Identify an Alarm Condition 54
7.2.2 Pause the Alarm Sound 55
7.2.3 Reset an Alarm 55
7.2.4 View Historical Alarms 56
7.2.5 Adjust the Alarm Sound Level 57
7.3 Physiological Alarms 57
7.3.1 High Pressure Alarm 58
7.3.2 Low Pressure Alarm 58
7.3.3 High Breath Rate Alarm 59
7.3.4 Low Breath Rate Alarm 59
7.3.5 High Minute Volume Alarm 60
7.3.6 Low Minute Volume Alarm 60
7.3.7 Rebreathing Alarm 61
7.3.8 Apnea Alarm 61
7.3.9 Disconnection Alarm 62
7.3.10 High EPAP Alarm 62
7.3.11 Low EPAP Alarm 63
7.3.12 High SpO_2 Alarm 63
7.3.13 Low SpO_2 Alarm 63
7.3.14 High Pulse Rate Alarm 64
7.3.15 Low Pulse Rate Alarm 64
7.3.16 Circuit Disconnection Alarm 65
7.3.17 Blocked Circuit Alarm 65
7.4 Technical Alarms 65
7.4.1 High Pressure Limitation Alarm 65
7.4.2 Power Fail Alarm 66
7.4.3 SpO_2 Sensor Failure / Disconnected Alarm 66
7.4.4 SpO_2 Artifact 67
7.4.5 Ambient Pressure Compensation Lost Alarm 67
7.4.6 Low Battery Alarm 67
7.4.7 Critically Low Battery Alarm 68
7.4.8 High Patient Air Temp. (High Patient Air Temperature) 68
7.4.9 Flow Sensor Failure 69
7.4.10 Internal Function Failure 69
8 Cleaning and Maintenance 70
8.1 Cleaning the Vivo 3 70
8.1.1 Clean the Main Unit Externally 70
8.1.2 Clean the Humidifier Water Chamber 71
8.1.3 Clean the Patient Circuit 71
8.2 Clean and Replace Patient Air Inlet Filters 71

8.3 Disposal 72
9 Technical Specifications 73
9.1 Ventilator Size and Weight 73
9.2 Power Supply 73
9.3 Operating Conditions 74
9.4 Emission and Immunity Declaration 75
9.4.1 Guidance and Manufacturer's Declaration – Electromagnetic Immunity 77
9.4.2 Guidance and Manufacturer's Declaration – Electromagnetic Emission 79
9.4.3 Recommended separation distances between portable and mobile RF communications equipment and the ventilator 79
9.4.4 Recommended separation distances between external power conductors and the ventilator 80
10 Accessories 81
10.1 Patient Circuits and Air Delivery Accessories 81
10.2 Power Accessories 83
10.3 Ventilator Filters and Detachable Parts 83
10.4 Monitoring Accessories 84
10.5 Other Accessories 86
Appendices 89
A FAA Compliance 89
Index 91

1 Introduction

WARNING!

Risk of Personal Injury

The Vivo 3 must only be used:

- For the intended treatment in accordance with this manual and with the instructions given by the responsible clinical personnel.
- In accordance with the operating conditions specified in this manual.
- In original and unmodified shape and only with compatible accessories.

Every other use may lead to risk of personal injury!

CAUTION!

Read this manual thoroughly so that you completely understand how the Vivo 3 is operated and maintained before taking it into use, to ensure correct usage, maximum performance and serviceability.
Non-professional caregivers (e.g. family members) should consult the medical equipment provider's respiratory therapist if they have any questions about the function, proper use, operation, service or maintenance of the Vivo 3.

WARNING!

The Vivo 3 is not designed for life support treatment:

- The Vivo 3 should not be used for life support treatment.
- The Vivo 3 shall only be used by patients with spontaneous breathing.
- The Vivo 3 should not be used for ventilator dependent patients.

1.1 Manufacturer Information

Legal Manufacturer

CE 2797

Postal Address
Breas Medical AB
Företagsvägen 1
SE-435 33 Mölnlycke Sweden

Web Address
www.breas.com

Email address
breas@breas.com

Phone
+46 31 868800, Order: +46 31 868820, Technical support: +46 31 868860

Fax
+46 31 868810

UK Sales and Support Information

UK Responsible Person: Breas Medical Ltd

Phone UK Head Office

+44 (0)1789293460

Fax UK Head Office

+44 (0)1789262470

General Enquiries and Ordering

orders@nippyventilator.com

Tech Support

techsupport@nippyventilator.com

Local Representatives Information

www.breas.com/contact-us/

1.2 What is the Vivo 3?

The Vivo 3 is an internally powered pressure ventilator capable of delivering invasive or non-invasive ventilatory support for the care of individuals who require long-term support from mechanical ventilation for respiratory insufficiency or respiratory failure, with or without obstructive sleep apnea.

1.2.1 Non Invasive Interfaces

The ventilator system can be used non-invasively with nasal mask, full/total face mask, and nasal pillow interfaces.

Nasal Cannula can be used with HFNT.

1.2.2 Mobility and Usage Environment

The ventilator system is classified as transit-operable and is intended to be used in homes, public spaces, institutions and hospitals.

The ventilator system is intended to be used together with portable applications such as wheelchairs, personal family vehicles, ground ambulances and civil aircraft (not helicopter). It is not intended for use during emergency transports.

1.2.3 Continuous Operation

The ventilator can be used for continuous operation up to 24 hours/day at least for 90 days without restarting.

1.2.4 Service Life

The expected service life of the Vivo 3 is 5 years or 20,000 hours.

1.2.5 Cybersecurity

The Vivo 1-2-3 devices do not have a network connection. It does not require a user account or password for its intended use.

The Vivo 1-2-3 has two interfaces:

- SD-Card, for log files and for devices with firmware at least v2.3.0 firmware upgrade (secured by only allowing binary files with matching signature generated by Breas.)
- USB port, for use by authorized service personnel to perform service and upgrade firmware. The USB port can also connect to the optional accessory iLink, a device used to connect to EveryWare over cellular 4G /LTE network. The iLink Instructions for use lists the prerequisites for using iLink.

The interfaces are only active when using the related functions.

The Vivo 1-2-3 software can be updated or reinstalled by authorized service personnel only.

1.3 Intended Use

Vivo 3 is intended to provide non-invasive or invasive ventilation for patients weighing over 10 kg (22 lbs) who require long-term support or mechanical ventilation for respiratory insufficiency or respiratory failure, with or without obstructive sleep apnea.

Vivo 3 is intended for spontaneously breathing patients.

1.4 Clinicial Benefits

The expected benefits for the ventilator non-dependent patient receiving long-term ventilatory support for respiratory insufficiency or failure are reduction of the associated clinical symptoms and the possibility of an increased quality of life.

1.5 Operation by Lay Users

Day-to-day caregivers, patients, relatives and other non-professional users may operate the Vivo 3 with the *Home mode* activated, after it has been set up according to the prescribed treatment. In Home mode, some settings and controls are locked or hidden.

The lay operator may also perform basic maintenance that doesn't require special equipment or a service environment.

Training

The lay operator shall be trained to basic knowledge of the Vivo 3 and in the specific operations they are assigned to perform. The training shall be based on this user manual and the responsible clinical personnel shall assess the level of training required for each lay operator. This manual shall be available for training and as reference when operating the Vivo 3.

1.6 Contraindications

The Vivo 3 is not a life-support ventilator and is contraindicated in patients who are unable to tolerate more than brief interruptions in ventilation.

If a patient has any of the following conditions, therapy with positive airway pressure may be contraindicated and the prescribing clinician shall decide if the benefit of ventilatory assistance overweighs the risks:

- Untreated pneumothorax
- Pneumomediastinum
- Inability to maintain a patent airway or adequately clear excessive respiratory secretions
- Severe acute systemic complications (shock, unstable arrhythmias, myocardial ischemia)
- Severe bullous lung disease
- Risk of vomiting
- Pathologically low blood pressure, especially if associated with intravascular volume depletion
- Cerebrospinal fluid leak, recent cranial surgery or trauma

The use of the Vivo 3 is contraindicated in an MRI environment.

Adverse Effects

If the patient experiences chest discomfort, pain, severe headache or shortness of breath while using the Vivo 3, a physician or responsible clinician should be contacted immediately.

The following side effects may occur during the course of therapy with the Vivo 3, patients are advised to report any new or changing adverse effects to their physician:

- Nasal, mouth or throat dryness
- Nosebleeds
- Abdominal bloating
- Ear or sinus discomfort
- Eye irritation
- Skin rashes

1.7 About this Manual

Firmware described in this manual: 2.3.4

1.7.1 Audience

This manual is intended for patients and other lay users operating the Vivo 3.



- Care providers, clinical personnel, physicians and others who require a working knowledge of the Vivo 3 will find additional information on settings and functions in the Clinician's Manual. The Clinician's manual should be of the same revision as the User's Manual.
- Service personnel may order the Service Manual that contains detailed technical information for maintenance, service, repair and disposal procedure. The Service manual's revision is independent of the User's Manual revision.

1.7.2 Icons in this Manual

In this manual, icons are used to highlight specific information. The meaning of each icon is explained in the table below.

Icon	Explanation
	**Warning!** Risk of death or personal injury.
	**Warning!** Risk of Cross-contamination.
	**Warning!** Risk of electric shock.
	**Warning!** Hot surface, risk of burns.
	**Warning!** Flammable material, risk of fire.
	**Caution!** Risk of equipment damage, loss of data, extra work, or unexpected results.
	MR Unsafe. The device should not enter a magnetic resonance (MR) environment, such as an MRI scanner room.
	**Note** Information that may be valuable but is not of critical importance, tips.
	**Reference** Reference to other manuals with additional information on a specific topic.

2 Safety Information

2.1 General Use — Warnings and Precautions

WARNING!

Risk of Personal Injury
The Vivo 3 is not designed for life support treatment:

- The Vivo 3 shall only be used by patients with spontaneous breathing.
- The Vivo 3 should not be used for ventilator dependent patients.

Risk of Faulty Treatment
If the patient is admitted to a hospital or is prescribed any other form of medical treatment, always inform the medical staff that the patient is on mechanical ventilation treatment.

The Vivo 3 shall not be used with nebulizers.

Risk of Insufficient Ventilation
Usage outside the specified operating conditions may cause reduced performance.
The Vivo 3 must only be used in accordance with the operating conditions specified in this manual.

Risk of Faulty Treatment
Do not use the Vivo 3 in the event of:

- Suspected damage to the device, including the occurrence of Internal Functional Failure alarms.
- Unexpected patient symptoms during treatment.
- Unexplainable or sudden changes of pressure, performance or sound during operation.
- Delivered air being abnormally hot or emitting an odor.

Contact your responsible care provider for an inspection.

Risk of Suffocation
Do not remove the protection for the heated patient circuit connector, the connector can be fitted with the protection in place.
If removed, children can choke or suffocate if swallowing it.

Risk of Asphyxia or Personal Injury
Using incorrect settings may cause personal injury or severe medical conditions.
The therapy settings shall be based on a physician's prescription. Changes to settings must be made by authorized clinical personnel only.

Risk of Faulty Treatment
The responsible organization should periodically reassess the settings of the therapy for effectiveness.

Risk of Faulty Treatment
Always prepare the Vivo 3 as described in this manual before use.

Risk of Unnoticed Critical Conditions

- All the physiological alarms of the Vivo 3 must be set at safe levels that will effectively warn the user of any risk.

  The alarm levels should be assessed considering the patient's treatment settings.
- Any change of treatment settings or change of components in the ventilation system may require readjustment of the alarm levels.
- The alarm sound level should be set to a clearly audible level. Setting the alarm sound level below that of the ambient sound level can impede recognition of alarm conditions.

Risk of Reduced Safety and Performance

Accessories that have not been verified to be compatible with the Vivo 3 might affect safety features and performance negatively.
Only use the Vivo with accessories that are compatible with the ventilator. Use of incompatible parts to connect the ventilator to the patient can result in degraded performance and change of pressure gradient.
Breas Medical has verified the compatibility between the Vivo 3 and the accessories listed in 10 *Accessories*, page 81.
The responsible organization must ensure the compatibility of the ventilator with all parts used to connect to the patient before starting the intended treatment. If incompatible accessories are used, Breas Medical has no responsibility for the safe and effective use of the Vivo 3.

WARNING!

Risk of Electric Shock

The Vivo 3 must only be used in original and unmodified shape and only with compatible accessories.
Inadequate use of device or accessories may cause loss of treatment or decreased performance.

WARNING!

Risk of Burns

Covering breathing tubes with a blanket or heating them with an overhead heater can affect the quality of the therapy or injure the patient.

CAUTION!

Proper placement and positioning of the patient interface is critical to the consistent operation of this equipment.

Handle the Vivo 3 with care.

Always use the therapeutic pressure setting, as individually determined with the configuration of the equipment and accessories.

Make sure to place and pack the device in a way that prevents unintentional start of the machine.

Due to the internal battery, the Vivo 3 may start if the Start/Stop button is pressed even without the mains being connected.

Do not use the Vivo 3 with nitric oxide, helium or helium mixtures. This may affect patient air flow and volume measurements.

If you suspect that the device has been mistreated, perform a functional check before taking it to use. A basic functional check can be performed as described in 5.5 *Performing Start-up Checks*, page 39. A complete functional check can be performed by an authorized service technician.

NOTE

Any serious incident that has occurred in relation to this device should be reported to the competent authority and the manufacturer.

2.2 Electricity — Warnings and Precautions

WARNING!

Risk of Electric Shock

- Do not operate the Vivo 3 if it has a damaged power cord, power supply or casing.
- To avoid electrical shock, only clean the Vivo 3 according to instructions in this manual. Do not soak or immerse the Vivo 3 into any fluids.
- Use the approved power supply units only.

  Use of unapproved power supply units may compromise the electrical isolation and lead to risk of electric shock.
- Do not use more than one multiple portable socket-outlet or extension cord.

  If a multiple portable socket-outlet is used, it must not be placed on the floor.
- The operator must not touch accessible contacts of connectors and the patient simultaneously.
- Nurse Call must only be connected to a safety extra low voltage system with an isolation from AC power (Mains) voltage which complies with the requirements of IEC 60601-1.

WARNING!

Risk of Faulty Treatment

Electromagnetic Interference may cause electrical equipment to malfunction.

- The aspects of electromagnetic compatibility must be considered.
  - The Vivo 3 should not be used adjacent to or stacked with other equipment; if adjacent or stacked use is necessary, the Vivo 3 should be observed to verify normal operation in that configuration.
  - Mobile or transportable radio transmitters may interfere with the Vivo 3.
  - Further guidance for safe installation of the ventilator can be found in the chapter about emission and immunity declaration.
- If a portable AC power supply is used, make sure that the voltage variations are within the operating limits of the Vivo 3.
- Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the Vivo 3, including cables specified. Otherwise, degradation of the performance of this equipment could result.

2.2.1 Electromagnetic Compatibility and Electrostatic Discharge (EMC and ESD)

Use of accessories, transducers and cables other than those specified or provided by Breas could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation.

Electromagnetic disturbance may impair the safety and performance of the Vivo 3. Unsteady indicated values for delivered volumes or pressures and the occurrence of alarm conditions without apparent cause may be an indication of loss of performance due to electromagnetic disturbances.

Measures for keeping electromagnetic field levels low should include but are not be limited to:

- Normal precautions with regard to relative humidity and conductive characteristics of clothing in order to minimize the build-up of electrostatic charges.
- Avoiding the use of radio emitting devices (e.g. cellular or cordless telephones, microwave ovens and high-frequency surgery apparatus) closer than 1 meter to the Vivo 3.
- Avoiding the use of known sources of Electromagnetic Interference, (e.g. RFID, diathermy equipment), in the presence of the Vivo 3.

  Please note some of these RF emitters may not be visible and the Vivo 3 can potentially be exposed to fields from these RF emitters without the user's awareness. If abnormal performance of the Vivo 3 is observed, and the RF emitters cannot be identified and removed, the Vivo 3 may need to be reoriented or relocated.

See the section 9.4 *Emission and Immunity Declaration*, page 75 for detailed information and further guidance for mitigating electromagnetic disturbance.

2.3 Environment — Warnings and Precautions

WARNING!

Risk of Faulty Treatment
If a room humidifier is used, place it at least 2 meters away from the Vivo 3.

Risk of Faulty Treatment

MR Do not use or store the Vivo 3 in a magnetic resonance (MR) environment. Use of the Vivo 3 in an MR environment may result in malfunction of the Vivo 3 and pose unacceptable risk to the patient, medical staff or other persons.

Risk of faulty Treatment
The performance of the Vivo 3 may deteriorate at altitudes or ambient temperatures outside the operation conditions specified in the chapter *Technical Specifications*.

- Do not use the ventilator while positioned in a warm place, such as direct sunlight or close to a radiator as this might lead to temperature outside the specifications.
- Do not use the ventilator in an hyperbaric chamber, as this would cause an ambient pressure outside the specifications.
- Do not use the ventilator immediately after storage or transport outside the recommended operating conditions.

WARNING!

Risk of Fire
Do not use the Vivo 3 in environments where explosive gases or flammable anesthetic agents present.

WARNING!

Risk of Electric Shock
Water on and in the device may cause an electric conductive path.
Do not expose the Vivo 3 to rain or snowfall.

CAUTION!

The ventilator, any accessories and all replaced parts, must be disposed of in accordance with the local environmental regulations regarding the disposal of used equipment and waste.

2.4 Patient Circuit — Warnings and Precautions

WARNING!

Risk of Insufficient Ventilation
The Vivo 3 ventilator is intended to be used with patient circuits with intentional leakage and compliant to ISO 17510. Recommended leak rate: 20 to 50 liters per minute at 10 cmH_2O.
Failure to use a mask or accessory that minimizes rebreathing of carbon dioxide or permits spontaneous breathing can cause asphyxiation.

Risk of Abnormal Exhalation Volume Measurement
Non-invasive ventilation: The exhaled volume of the patient can differ from the measured exhaled volume due to unintentional leaks around the mask.
Invasive ventilation: The exhaled volume of the patient can differ from the measured exhaled volume due to unintentional leaks around the cannula.
For correct measurements, minimize unintentional leaks.

Risk of Reduced Performance
Filters and patient connected parts must be replaced regularly to ensure correct function of the Vivo 3.

Risk of Insufficient Ventilation
Before use:

- Make sure that the patient circuit and joined parts are undamaged and correctly connected, in order to avoid unwanted leakage.
- The leakage port of the patient circuit or patient interface prevents rebreathing by flushing the exhaled air. It should be located as near the patient interface as possible (this is even more important for treatments with low pressure). Make sure that it is not blocked or obstructed.
- The Vivo 3 should be turned on and the function of the leakage port should be checked before use.

Risk of Insufficient Ventilation
Incompatible patient circuits may come loose.
To prevent disconnection of the patient circuit or patient circuit system during use, especially during ambulatory use, only patient circuits in compliance with ISO 5367 or ISO 80601 - 2 - 74.

Risk of Suffocation
Periodically check for moisture in the patient circuit.
When present, remove the moisture. Before attempting to dry the circuit, disconnect it from the Vivo 3 to ensure no water flows back into the Vivo 3.
The frequency at which these checks must be performed will depend on the patient's condition and the device used. The responsible caregiver should assess this on an individual basis in accordance with the patient's needs.

Risk of Suffocation
If the patient needs assistance to remove the patient interface, the patient shall not be left alone. This is to avoid the risk of re-breathing of CO_2 in case of accidental ventilator failure.
Do not breathe through the connected patient circuit unless the ventilator is turned on and operating properly.

Risk of Suffocation
If the patient is using a full face mask (covering mouth and nose), the mask must be equipped with an anti-asphyxia valve.

Risk of Suffocation or Barotrauma
In HFNT mode, do not use a sealed patient interface.

Risk of Constriction
Entanglement with cables or tubing constricting airways may cause asphyxiation.
Do not leave long lengths of air tubing or cables around the top of the bed. They could twist around the patient's head or neck while sleeping.

Risk of Insufficient Ventilation
The use of equipment such as endotracheal tubes, oral/nasal tubes, adaptors etc. with small inner diameters or high resistance filters, humidifiers etc. increase the resistance in the patient circuit which may interfere with the operation of the patient disconnect function. It may also interfere with the device trigger function.

WARNING!

Risk of Electric Shock
Do not use antistatic or electrically conductive hoses or tubing with the ventilator breathing system.

WARNING!

Risk of Cross-Contamination
Patient circuits might get contaminated by exhaled gases. To avoid cross-contamination, always use a new patient circuit when the Vivo 3 is to be used by a new patient.

NOTE

For patient interfaces and accessories, always follow the manufacturer's instructions.

2.5 Filter Usage — Warnings and Precautions

Low resistance bacteria filter, if used, should be replaced every 24 hours.

WARNING!

Risk of Overheating

Replace or clean the air inlet filters as specified in the *Maintenance* chapter.
Using old or clogged filters may cause the Vivo 3 to operate at higher temperatures than intended.
When operating the Vivo 3, make sure that the air inlet and filters are not obstructed or occluded.

Risk of Insufficient Ventilation

Do not use high resistance bacteria filter at the air outlet of the Vivo 3. High resistance bacteria filters placed between the air outlet and the patient interface may interfere with the operation of the patient disconnect function. It may also interfere with the device trigger function.

Risk of Insufficient Ventilation

Humidification can increase the resistance of breathing system filters and the operator must monitor the breathing system filter frequently for increased resistance and blockage to ensure the delivery of the therapeutic pressure.
Increased resistance may interfere with the operation of the patient disconnected function.

WARNING!

Risk of Cross-Contamination

- If assigning the Vivo 3 to a new patient, always replace the filters.
- Always use the Vivo 3 with patient air inlet filters installed.

Risk of Cross-Contamination

Deep tissue or mucosal contact with infectious agents may cause infections.
If the Vivo 3 is to be used as a multiple patient device, a low resistance bacteria filter shall be used between the air outlet and the patient circuit, for preventing patient cross-contamination. Reuse of bacteria filter, patient circuit or mask may expose patients to contagious agents.

2.6 Invasive Use — Warnings and Precautions

WARNING!

Risk of Faulty Treatment

If using the Vivo 3 for invasive ventilation, humidification is recommended. Make sure to use an external heated humidifier that complies to invasive use according to ISO 80601-2-74.
The attachable humidifier is not intended for invasive use.

Risk of Unnoticed Critical Conditions

When using the Vivo 3 invasively, physiological alarms, such as the alarms listed below, must be carefully set, to ensure safe use.

- Low volume alarm
- Low breath rate alarm
- Low pressure alarm
- Disconnection alarm

Risk of Rebreathing
Be aware of risks for rebreathing exhaled gases:

- A blocked, clogged or in other ways occluded leakage port causes exhaled gases to be rebreathed.
- With low pressures the risk of rebreathing increases.
- I:E (inspiration time : expiration time) ratios close to 1:1 increase the risk of rebreathing as the expiratory time allowing exhaled gases to be cleared from the circuit before the next breath is decreased.

The Vivo 3 is equipped with a low leakage alarm. However, the alarm is not a substitute for operator vigilance for ensuring that the leakage ports remains clear at all times. Periodically check the leakage ports during therapy.
See the section 2.4 *Patient Circuit — Warnings and Precautions*, page 14 for mitigating the risks of rebreathing.

2.7 Humidification and Heating — Warnings and Precautions

WARNING!

Risk of Personal Injury
The attachable humidifier shall not be used with nitric oxide. Such use might cause the humidifier to not function correctly, causing serious deterioration of health.

Risk of Suffocation
When the attachable humidifier is installed, the Vivo 3 must be located below the patient and on a flat surface. This is to prevent accidental spillage or condensate flowing down the patient tube and into the patient's mask. Extra cautions should be taken for patients who are unable to guard their airways or cannot pull off the mask.

Risk of Suffocation or Loss of Ventilation
If using an external humidifier, it shall be placed below both the patient and the Vivo 3. This is to prevent accidental spillage or condensate flowing down to the ventilator or down the patient tube and into the patient's mask. Extra cautions should be taken for patients who are unable to guard their airways or cannot pull off the mask.

The use of external humidifier may require readjustment of the low-pressure alarm.

Risk of Suffocation
Installation of a water trap may be required if the condensation is extensive in the patient circuit when using a heated humidifier.
The water trap prevents condensed water in the patient circuit from reaching the patient airways and causing personal injury.

Risk of Suffocation
Do not use the attachable humidifier during mobile use. Due to movements, water from the humidifier may flow through the patient circuit and cause suffocation.

WARNING!

Risk of Electric Shock
Water spillage may cause electric conductive paths resulting in electric shocks.

- Do not use the attachable humidifier during mobile use.
- Do not transport the ventilator with water in the attachable humidifier.
- The attachable humidifier must be detached before filling. Do not fill above the *Maximum Water Level* indication on the water chamber.

WARNING!

Risk of Burns
After using the ventilator, wait one minute before you open the water chamber since it can get hot under certain conditions (for example if the humidifier runs out of water).

2.8 Cleaning and Maintenance — Warning and Precautions

This manual contains instructions for cleaning and maintenance that can be carried out by the care provider or users with physical ability and working knowledge of the system.

WARNING!

Risk of Faulty Treatment
Service and Maintenance of the Vivo 3 shall not be performed when the Vivo 3 is in use.

WARNING!

Risk of Electric Shock
Cleaning with excessive water or opening the device's casing without certified training may cause electric shocks.
The Vivo 3 should be regularly cleaned and maintained in accordance with this operating manual.

Risk of Electric Shock
High voltage contact may cause cardiac arrhythmia.
Repairs and modifications must be carried out by authorized technicians only and in accordance with instructions from Breas Medical

- The Vivo 3 must not be opened, repaired or modified by unauthorized personnel or interconnected with incompatible equipment. If subjected to unauthorized modifications or operations, Breas Medical is no longer responsible for the device's performance and safety and all warranties will become invalid.

CAUTION!

Do not attempt to autoclave or sterilize the Vivo 3.

2.9 Oxygen Usage — Warning and Precautions

When using the Vivo 3 with oxygen, always follow the oxygen provider's instructions and use only medical grade oxygen complying with local regulations.

WARNING!



As this medical device uses an alternative small-bore connector design different from those specified in the ISO80369 series, there is a possibility that a misconnection can occur between this medical device and a medical device using a different alternative small-bore connector, which can result in a hazardous situation causing harm to the patient. Special measures need to be taken by the user to mitigate these reasonable foreseeable risks.



Do not use a humidifier between the oxygen source and the ventilator, in order to humidify the oxygen flow.
If humidification is required, use the attachable humidifier or an external humidifier after the patient air outlet.



Risk of faulty Treatment
At a fixed flow rate of supplemental oxygen flow, the inhaled oxygen concentration will vary, depending on the pressure or flow delivered, the patient interface and leak rate.
To monitor the oxygen concentration, use an external patient air oxygen concentration monitoring unit which complies with ISO 80601-2-55 and is equipped with a high oxygen level alarm.



Risk of faulty Treatment
The responsible organization shall ensure that the oxygen source is compatible with the rated range of pressure, flowrate and oxygen concentration for this device, see *Supplemental Oxygen Supply Requirements*, page 49.
Failure to use a compatible oxygen source may result in serious deterioration of health.

WARNING!

Risk of Fire

The presence of oxygen makes materials more flammable and speeds up the combustion.

- Turn the oxygen supply off when the ventilator is not delivering treatment. Oxygen may otherwise accumulate in the enclosure and increase the risk of fire.
- When the oxygen supply is turned on, never leave the patient interface on textiles like bed coverings or chair cushions. The oxygen enrichment will make the materials more flammable.
- Use only water-based lotions or salves that are oxygen-compatible before and during oxygen therapy. Never use petroleum-based or oil-based lotions or salves, to avoid the risk of fire and burns.
- Covering breathing tubes with a blanket or heating them with an overhead heater can affect the quality of the therapy or injure the patient.
- Ventilate the room adequately.
- Do not smoke in a room where oxygen is being used, it may result in facial burns or death. If the patient intends to smoke: turn the equipment off, remove the patient interface and leave the room where the equipment is located before smoking. If unable to leave the room, wait 10 minutes after you have turned the equipment off.
- Naked light bulbs, open flames and other sources of ignition must be kept a minimum of 2 meters (6 feet) away from the oxygen cylinder or any oxygen carrying parts and accessories.
- Do not use aerosols or solvents close to the oxygen supply, even when the oxygen supply is shut off.
- Do not lubricate fittings, connections, tubing, or other accessories of the equipment as the lubrication might be flammable in combination with the oxygen rich gas flow.

CAUTION!

Supplemental oxygen is added before the volume measurement sensor and thereby included in the measurements. However, the oxygen concentration still has influence on the volume measurement of the delivered air.
This measurement is based on a normal oxygen concentration of 21%. If the oxygen concentration is higher, the actual inspired volume will deviate from the monitored volume as follows:

- 40% oxygen concentration: -2.5% deviation
- 60% oxygen concentration: -5% deviation
- 80% oxygen concentration: -7.5% deviation

2.10 Mobile Use — Warning and Precautions

This section applies if using the Vivo 3 during transit, for example on a wheel chair or in a car.

WARNING!



Risk of Electric Shock
If connecting the Vivo 3 to an external DC power source, always use the DC DC Power Supply accessory. Connecting directly to an external DC power source may compromise the electrical isolation and cause an electric shock.
If connecting the Vivo 3 to a portable AC power generator, make sure its voltage variations are within the operating limits of the Vivo 3.

CAUTION!



During mobile use, protect the device with either of the Protective Cover accessory or the Lightweight Mobility Bag accessory.
Do not use the Vivo 3 while in a carry bag.

3 Product Description


This section describes the main Vivo 3 medical electric equipment.

For information about accessories and user replaceable spare parts., see .10 *Accessories*, page 81 or http://www.breas.com.

3.1 Overview


This overview shows the ventilator with operator detachable parts connected. For detailed information about connecting parts and accessories, see 5 *Prepare the Vivo 3 for Use*, page 37 and 6.6 *Using Accessories*, page 44.

1. Ventilator unit
2. Filter holder with air inlet filter
3. Power Supply
4. Patient Circuit
5. Air bypass unit. May be replaced by the attachable humidifier (accessory)




3.2 Ventilator Front

This section describes the front panel.



1. **Screen**, displays pages with information, settings and commands.
2. **Directional buttons**, moves between and selects objects on the current page.
3. **On/Off button**, starts /stops treatment or turns the ventilator On/Off.
4. **Navigation buttons**, select pages according to the corresponding label on the display. The navigation buttons can temporarily be designated functions for replying to questions or requests from event windows.
5. **Audio pause button**, pauses the alarm sound.
6. **Audio Pause LED** , lights yellow when the alarm sound is paused.
7. **Alarm LED**, flashes during active alarms.

3.3 Ventilator Back



1. **Power port**, for connecting the power supply.
2. **Oxygen port**, for low pressure/bleed-in oxygen
3. **Patient air outlet**, connection for patient circuit
4. **Patient air inlet**, with filter holder
5. **Air bypass module**, directs the patient air flow. Removed if the attachable humidifier is used.
6. **USB port**, for data transfer to a PC.
7. **SD Card port**, for copying records and logs to a PC.
8. **Communication port**, for connection of accessories.

The ventilator can be carried by hand using the handle.

3.4 Power Management

WARNING!

Do not connect the ventilator directly to the battery of a wheelchair or any similar external power source, as this can compromise the ventilator performance and result in degradation of the health of the patient.
For external DC power other than the Breas XPAC, always use the DC/DC power supply adapter (accessory).

The Vivo 3 has a power management system that automatically selects the best available power source, according to the priority list below:

1. Mains power via Breas Mains power supply
2. External DC via the Breas XPAC, or the Breas DC/DC power supply adapter
3. Internal battery

The power source is indicated by a symbol at the top of the display.

Power Source	Icon
Mains	
External DC	
Internal battery	

For connecting the Vivo 3 to both mains and external DC at the same time, the accessory *Power supply Y-cable* is required.

If a power source fails, the Vivo 3 will switch to the next source in priority and show a message on the display. If all available sources fail, the *Power Failure* alarm is given and the Vivo 3 shuts down.

Charging the Internal Battery

The internal battery is automatically charged when the Vivo 3 is connected to external power. The state of charge is indicated by a symbol on the display. The green part of the battery shows the state of charge.

Battery State of Charge	Icon
High	
Charging	

Preparing the Battery for Storage


If storing the Vivo 3, longer than 1 month, make sure that the internal battery is about half charged before storing, in order to maintain maximum battery capacity. Optimal storage temperature is 5 to 30°C (41 to 86°F).

3.5 Menus

3.5.1 Use the Menu

The menu consists of four function sections:

- Monitor
- Setup
- Alarm
- Others



a. Navigation buttons

b. Active page indicator

c. Directional buttons

d. Page number indicator (for functions with several pages)


Selecting the Section to Display

1 Press the navigation button for the requested function page.

⇒The page is now displayed.

2 For functions with several pages grouped together, press the navigation button again to browse the pages.

3 For pages with menus, settings or additional information, press the Up and Down arrow buttons to select an item on the page.





Select an Item on a Page

This procedure describes how to navigate between selectable items. Read-only information cannot be selected.

Selectable Items


- Sub pages (indicated by an arrow to the right of the item text)
- Settings
- Commands

1 Press the Up or Down arrow button to select the first selectable item on the page.

⇒The selected item is highlighted.

2 To select another item on the list, press the Up or Down arrow until it is highlighted.

Note that items with read only information cannot be selected.



Enter a Sub Page

1 Select the sub page (indicated by an arrow to the right of the item text) using the Up or Down arrow button.

2 Enter the sub page by pressing the right arrow button.

⇒ The sub page is now displayed.

3 Leave the sub page using the left arrow button



Change a Setting

1 Select the setting (indicated by an arrow to the right of the item text) using the Up or Down arrow button.

2 Change the setting using the right or left arrow buttons to select between the predefined values.

⇒ The currently displayed value will remain selected when leaving the page.

3 Leave the sub page using the navigation buttons.



Execute a Command


1 To execute a command, select it and press the right arrow button.

⇒ The command execution is started. Additional actions related to the command might be requested in event windows.



3.5.2 The Monitor Page

The monitor page displays the treatment data. It consists of a bar graph illustrating the current pressure and text area displaying current monitored values in text. See also 6.5.1 *Treatment Values Monitored by the Vivo 3* , page 44.



Bar graph legend

1. Current pressure
2. Low pressure alarm level
3. Set inspiration pressure level
4. High pressure alarm level


Monitor page with HFNT

When HFNT is active, flow is displayed instead of pressure.



3.5.3 The Setup Pages

The setup pages contain settings related to the treatment.



Selecting a profile

If profiles with preset treatment settings have been created, the profiles are selected from the Setup section.

1 Select Profiles and press the Down or Up arrow button to select one of the profiles.

2 Press Left arrow button to select a profile. The currently active profile is indicated by a checkmark.

For detailed information about each setting, see 4 *Treatment Functions and Settings*, page 35.

3.5.4 The Alarm Pages

The alarm pages contain alarm settings and an alarm history list.

For information about the alarms and their possible settings, see 7 *Alarms*, page 53.

3.5.5 The Others Pages

The Others pages contain non-clinical settings and information.

3.5.5.1 Device Settings

This section describes device settings that don't affect the ventilating function of the Vivo 3.

Setting	Description
Language	**Value range:** The available languages are listed with their native names accompanied by a flag representing a country where the language is spoken. **Default value:** English (The language selection menu is displayed when the Vivo 3 is started for the first time.)
Confirm Start/Stop	On — The user need to confirm that the treatment is to be stopped. Off — The treatment is started and stopped directly when pressing the start/stop button on the machine. Default value: On
Auto Start	On — The treatment starts automatically when an inhalation is detected. Off — The treatment is started by pressing the start/stop button on the machine. Default value: Off Auto start is a convenience function. It may not be compatible with all types of masks.
Pressure Unit	**Value range:** • cmH_2O • mbar • hPa **Default value:** cmH_2O
Display Light	**Value range:** • On — Always lit at the selected intensity • Auto — Always lit, with automatic adjustment of the intensity with regards to the ambient light. • Delayed — The display is dimmed after about 30 seconds (time depends on usage mode and power source). If any button is pressed or any alarm occurs, the display is lit again. • Night — The display and button backlights are turned off. If any button is pressed or any alarm occurs, the display and back lights are lit again. **Default value:** Delayed
Light Intensity	**Value range:** 1 to 5 **Default value:** 5

3.5.5.2 Pre-use Test

The pre-use test checks the compatibility of the circuit, including connected accessories, by measuring the resistance and compliance. If any changes are made to the patient circuit configuration, it is recommended to perform a new pre-use test.

If the patient circuit is set to *Custom* during setup, the measured values will be used for optimizing the treatment.

The pre-use test is performed using a guide in the graphical user interface, see 5.6 *Performing a Pre-Use Test*, page 39.

3.5.5.3 Device Information

Information	Description
Product name	
Operating time	The total number of hours the ventilator has been running in operating mode.
Battery Percentage	Battery charge level
Battery Time Left	Estimated remaining usage time on battery power
Software module versions	Technical information for service personnel

3.5.5.4 Device Memory

Command	Description
Save Memory Data on Card	Copies the memory data from the internal memory to the memory card.
Erase Memory Data	Erases memory data from the internal memory
Remove SD-card	Remove the SD-card from the ventilator. (It must be ejected manually, though.)

3.5.5.5 Clock

Setting	Description
Time	Sets the time for the Vivo 3. The time is used for logs and reports.
Date	Sets the date for the Vivo 3. The date is used for logs and reports.
Time Format	Select whether to use 12 hr clock or 24 hr clock.
Date Format	Select the date format to use.
Alarm Clock	Activates an alarm clock on the Vivo 3. One short signal every 5 seconds.
Alarm Clock Time	Sets the alarm time for the alarm clock.
Alarm Clock Volume	Sets the volume for the alarm clock.

NOTE

The alarm clock works only when the ventilator is running on mains.

3.5.5.6 Compliance Data

Setting / information	Description
Period	Period of time to view data.
Min daily use	The minimum daily use, in hours, to reach compliance per day.
Reset Compliance Data	Reset of data.
Start Date	Start date for treatment.
Total Usage Hours	The total number of hours the ventilator has been running in operating mode during the download period.
Total Days	The total number of days in the download period.
Days with Usage	The number of days in the download period where the ventilator has been running in operating mode (all day, or part of the day).
Average Usage Hours	The average number of hours per day the ventilator has been running in operating mode. Only days where the ventilator has been running in operating mode are part of the value (days without treatment are excluded in the calculation).
Days Compliant	Number and percentage of days when minimum daily use has been reached.
AHI	Average Apnea Hypopnea Index for selected period.*
AI	Average Apnea Index for selected period.*
HI	Average Hypopnea Index for selected period.*
IPAP	Average IPAP for selected period.
EPAP	Average EPAP for selected period.
Leakage	Average Total Leak for selected period.
Vte	Average Vte for selected period.
MVe	Average Mve for selected period.
Total Rate	Average Total Respiratory rate for selected period.
% Spont Breaths	Average % Spontaneous breaths for selected period.
Insp Time	Average Inspiratory Time for selected period.
I:E	Average I:E Ratio for selected period.
*) The displayed value is rounded to the nearest whole number.	

3.6 Symbols on the Vivo 3

Symbol	Description
	Manufacturer information
	Nemko certifikation mark (NRTL/SCC accredited)

Symbol	Description
IP22	Degree of protection provided by enclosure. See 9.3 *Operating Conditions*, page 74 for explanation.
Rx Only	Caution: U.S. Federal law restricts this device to sale by or on the order of a licensed healthcare practitioner. (Symbol only applicable in U.S.)
	Read user instructions
	RTCA/DO-160 G categorization. Category M This category is defined for equipment and interconnected wiring located in areas where apertures are electro-magnetically significant and not directly in view of radio reciever's antenna. This category may be suitable for equipment and associated inteconnecting wiring located in the passsenger cabin or in the cockpit of a transport aircraft.
CE 2797	Meets all requirements for CE marking according to relevant European health, safety and environmental protection legislation.
UK CA UK CA 0086	Meets all requirements for UKCA marking according to applicable United Kingdom health, safety and environmental protection legislation.
	Internal Battery Information
	Read the “Disposal” section for information about recycling and disposal, see 72.
REF	Product number
SN	Serial number
MD	This product is a Medical Device.
	Date of manufacture
O_2	Oxygen connection port. Max 30 l/min and 100 kPa.
	Power connection port. Use approved power supplies only.
	USB port
	SD memory card port
I/O	I/O port for accessory box/SpO_2

Additional Symbols on Parts and Accessories

This section describes additional symbols for Vivo 3 detachable parts and accessories from Breas Medical. Each item, or its package, have the symbols that applies for the specific accessory.

Symbol	Description
	Caution, hot surface
	Caution symbol, read the Accessory's instructions for more information.
	Keep away from rain.
	Single patient multiple use.
	Meets all requirements for CE marking according to relevant European health, safety and environmental protection legislation.
	Meets all requirements for UKCA marking according to applicable United Kingdom health, safety and environmental protection legislation.
	Applied part, type BF Electrically connected to the patient but not directly to the heart.
	IEC protection Class II: Double insulated equipment.
	Remote alarm port
	Alarm nurse call port
	Effort belt port (not used)
	AC power port
	DC power port
SP O2	SPO_2 port

3.7 Symbols on the Display



1. SD card inserted and working
2. SD card failure
3. Nurse call connected
4. Remote alarm connected
5. SpO2 connected
6. Humidifier activated
7. Heated circuit activated
8. “Clock radio wake up” activated
9. All alarms shut off
10. Active profile
11. Powered by mains.
12. Powered by external battery (no icon = powered by internal battery)
13. Battery status (red — yellow — green). A flash indicates charging.

4 Treatment Functions and Settings

This chapter describes the modes, settings and parameters that controls the ventilation of the Vivo 3.

All stated tolerances includes measurement uncertainty. The accuracies have been tested with all allowed configurations. Stated tolerances only disclose the maximum tolerance. If a parameter's tolerance is described with both absolute and relative measures, the greater one applies.

Lay operators must only use the User Manual, not the Clinician's Manual.

4.1 Treatment Settings

This section describes settings and parameters that affects the ventilating function of the Vivo 3. On the machine, these settings are in the **Settings** menu.

4.1.1 Fall Time

Fall Time is a comfort setting that might promote more natural breathing for some patients. A longer fall time might also reduce mask movements caused by rapid pressure transitions.

Fall Time controls the nominal speed of the depressurization during the exhalation phase and targets a linear curve of the pressure decay. The fall time is set as a percentage of the preceding inspiratory time.

Limitations for the *Fall Time* setting:

- If breath rate is set for any of the modes PSV, PCV+A or S/T, fall time is not applied to ventilator initiated breaths.
- The fall time cannot exceed 70% of the individual breath's maximum expiratory time.
- The maximum fall time is 3 seconds.
- The start of an inspiratory phase (spontaneous or timed) have precedence over any remaining fall time.

Fall Time can also be set to *Off*, which will give a minimal pressure decay time, based on the difference between IPAP and EPAP values.

Min	Max	Default	Resolution	Tolerance
Off, 10%	90%	Off	10%	—

4.1.2 Ramp Up

The ramp up setting defines a ramp time for increasing the airway pressures.

At the beginning of Ramp Up time, IPAP will start at 2 cmH_2O above the configured Ramp Pressure. If the Ramp Pressure was set below EPAP, the EPAP pressure will start at the configured Ramp Pressure.

Min	Max	Default	Resolution	Tolerance
5 minutes Off	60 minutes	Off	5 minutes	5 %

4.1.3 Ramp Down

The ramp Down setting defines a ramp time for decreasing the airway pressures.

At the end of Ramp Down time, IPAP will end at 2 cmH_2O above the configured Ramp Pressure. If the Ramp Pressure was set below EPAP, the EPAP pressure will end at the configured Ramp Pressure.

Min	Max	Default	Resolution	Tolerance
5 minutes Off	60 minutes	Off	5 minutes	5 %

4.1.4 Humidifier

This setting defines whether the attachable humidifier shall be used. If using an external humidifier, this setting shall be *Off*.

WARNING!

Read the section 2.7 *Humidification and Heating — Warnings and Precautions*, page 17 before activating the humidifier, to make sure all conditions are met and considered.

Min	Max	Default
Off	On	Off

4.1.5 Humidifier Level

This setting defines the level of humidification, if *Humidifier* is set to *On*.

Min	Max	Default
1	5	1

4.1.6 Heated Circuit

This setting defines whether a heated patient circuit shall be used as patient circuit.

WARNING!

Read the section 2.7 *Humidification and Heating — Warnings and Precautions*, page 17 before activating the heated circuit, to make sure all conditions are met and considered.

Min	Max	Default
Off	On	Off

4.1.7 Heated Circuit Level

This setting defines the heating level, if *Heated Circuit* is set to *On*.

Min	Max	Default
1	5	1

5 Prepare the Vivo 3 for Use


5.1 Checking the Vivo 3 before First Use

Before using the Vivo 3, perform the following checks.

1 Ensure that the equipment is in good condition.

2 If stored more than 1 month, connect the Vivo 3 to the power supply to recharge the internal battery.

3 Check that the air inlet filters are installed.



5.2 Placing the Vivo 3


WARNING!


Read the chapter 2.3 *Environment — Warnings and Precautions*, page 13 carefully to make sure all conditions are met and considered.

1 Place the Vivo 3 on a solid, flat surface. The Vivo 3 should be placed lower than the patient in order to prevent the device from falling on the patient, as well as preventing condensed water from reaching the patient.

2 Make sure that nothing can block the patient air inlet at the back of the Vivo 3, such as a curtain etc.

3 Make sure that the controls are accessible for the operator.





5.3 Connecting the Vivo 3 to Power Supply

This chapter describes how to connect a Breas power supply.

WARNING!

Read the chapter 2.2 *Electricity — Warnings and Precautions*, page 12 carefully to make sure all conditions are considered and met.

CAUTION!


Isolation from external power sources is provided by the approved AC/DC and DC/DC adapters. The ventilator must only be powered by the approved accessories, see 10.2 *Power Accessories*, page 83.

1 Plug the power supply´s connector into the power port at the back of the Vivo 3 (a).

2 Turn the connector clockwise 90 degrees (b).

3

- If using the AC/DC power supply, connect the AC plug to the mains supply outlet.
- If using DC/DC adapter, connect the DC plug to the external DC source.
- If using the XPAC, connect the battery cable to the XPAC main unit.



4 Make sure that the power outlet is not blocked, so that the cord can be unplugged without difficulties.

5.4 Connecting the Patient Circuit

WARNING!

Read the chapter 2.4 *Patient Circuit — Warnings and Precautions*, page 14 Carefully to make sure all conditions are considered and met.

The Vivo 3 is intended to be used with leakage circuits only. Recommended leak rate: 20 to 50 liters per minute at 10 cmH_2O.

1 Check that the circuit is clean and undamaged.

2 Connect the tube to the air outlet.

If having a heated circuit, make sure to connect the end with the heating plug to the to the air outlet. Refer to 5.4.1 *Connect the Heated Patient Circuit*, page 38 and 5.4.2 *Disconnect the Heated Patient Circuit from the Ventilator*, page 39.

3 Check whether the patient interface has an integrated leakage port.

- If yes, connect the patient circuit to the patient interface.
- If no, first connect a leakage port to the patient circuit, then connect it to the patient interface.


A bacterial filter can be attached on the ventilator outlet connector.

5.4.1 Connect the Heated Patient Circuit

1 Check the patient circuit for damages.

2 Go to Setup to define which circuit is being used. See .

3 Connect the ventilator cuff to the air outlet of the ventilator. A clicking sound is heard when the latches are fitted correctly.



4 If applicable, perform a pre-use test on the ventilator. See 3.5.5.2 *Pre-use Test*, page 30.

5 Connect the patient interface cuff to the patient interface.

6 On the ventilator, activate the circuit heating.

5.4.2 Disconnect the Heated Patient Circuit from the Ventilator

1 Disconnect from the ventilator:
Press the latches and pull the circuit off from the ventilator. Do not pull by the tube.

2 Disconnect from the patient interface:

Hold by the cuff and pull the circuit off from the patient interface. Do not pull by the tube.

5.5 Performing Start-up Checks

This procedure may be performed as a functional pre-use check at start-up.

Procedure

Prerequisites

- The Vivo 3 shall be connected to the power supply.

1

Start the treatment.

- If *Confirm Start/Stop* is *Off*, press **On/Off** button shortly.
- If *Confirm Start/Stop* is *On*, press and hold the **On/Off** button until the *Starting Treatment* progress bar is filled.

For information about setting the *Confirm Start/Stop* options, see 3.5.5.1 *Device Settings*, page 29 and 6.2 *Start the Treatment*, page 41.

.

2 When starting treatment, an automatic alarm signal test is performed. Check that the test is performed successfully, this is indicated by:

- A short beep indicating functional audio signaling.
- The alarm LED first lights yellow, then red, indicating functional visual signaling.
- The audio pause LED lights yellow.
- In about a second, both LEDs are turned off.

If the test fails, do not use the Vivo 3. Contact your supplier of the Vivo 3 for a technical check.

3 Disconnect the power cord for more than 5 seconds.

Check that the device switches to the internal battery and that the information message *Switched to Internal Battery* is shown on the screen.

If not, contact your supplier of the Vivo 3.

4 Reconnect the power cord.

Check that the Vivo 3 switches to the mains supply.

If not, contact your supplier of the Vivo 3.

5.6 Performing a Pre-Use Test

The pre-use test checks the compatibility of the patient circuit, including connected accessories, by measuring the resistance and compliance. If any changes are made to the patient circuit configuration, it is recommended to perform a new pre-use test.

When performing the pre-use test, the patient circuit and accessories up to (but not including) the leak port shall be connected. The patient shall not be connected during the pre-use test.

1 On the **Others** menu, select **Pre-use Test** and then press **OK** to confirm the start of the pre-use test.

2 Follow the instructions on the display and review the results at the end of the test.

5.6.1 Actions At Pre-Use Test Failure

At the end of the pre-use test the individual results for leakage, resistance and compliance are shown.

Failure Due To Incorrect Leakage

Indication: **Leakage: Fail**

1 Check the ventilator set-up (circuit, filters, humidifier, etc.) for leakage.

2 Ensure that all connectors are tightly fitting.

3 Run the pre-use test again.

4 Replace the circuit if the test is failed repeatedly.

Failure Due To Incorrect Resistance or Compliance

Indication: **Resistance: Fail** or **Compliance: Fail**

1 Check the ventilator set-up (circuit, filters, humidifier, etc.) for blockage or pinched tubing.

2 Run the pre-use test again.

If the pre-use test is continually failed due to resistance or compliance, it is permitted to use the ventilator but be aware that the pressure (resistance) or volume (compliance) delivered to the patient may not meet with the specified accuracy.

The ventilator will apply the default values to compensate for circuit resistance and compliance. These values will deviate from the values for the circuit in use.

Ensure that the delivered ventilation is closely monitored.

6 How to Use the Vivo 3

For a summary description of the means of initiating and terminating the inspiratory phase in each mode of the ventilator, please refer to .


6.1 Switch On the Vivo 3

- **Switching on an unpowered ventilator**

  Connect the power supply.

  ⇒The ventilator starts up and enters standby mode.

  If the power supply is connected during the *Power Failure* alarm, the ventilator will start the treatment directly instead of entering standby mode.
- **Switch on a powered ventilator**

  Press the **On/Off** button. 

  ⇒The ventilator starts up and enters standby mode.

6.2 Start the Treatment



WARNING!

Risk of Asphyxia or Personal Injury
Using incorrect settings may cause personal injury or severe medical conditions.
The therapy settings shall be based on a physician's prescription. Changes to settings must be made by authorized clinical personnel only.

It is possible to choose how the machine will start and stop treatment. For setting the Treatment start method, see 3.5.5.1 *Device Settings*, page 29

Auto Start is set to On

1 Make sure that the Vivo 3 is powered on.


2 With the patient interface fitted, make an inhalation and the Vivo 3 will start the treatment.

3 During treatment start, an alarm function test is performed, see *Alarm Function Test*, page 42. If the test fails, take the Vivo 3 out of use and contact your supplier of the Vivo 3 .

Confirm Start/Stop is set to Off

1 Make sure that the Vivo 3 is powered on.

2 Press the **On/Off** button.




3 During treatment start, an alarm function test is performed, see *Alarm Function Test*, page 42. If the test fails, take the Vivo 3 out of use and contact your supplier of the Vivo 3 .

Confirm Start/Stop is set to On

1 Make sure that the Vivo 3 is powered on.

2 Press and hold the **On/Off** button until the *Starting Treatment* progress bar is filled.

If the function test beep absents, take the Vivo 3 out of use and contact your supplier of the Vivo 3 .





3 During treatment start, an alarm function test is performed, see *Alarm Function Test*, page 42. If the test fails, take the Vivo 3 out of use and contact your supplier of the Vivo 3 .

Alarm Function Test

During treatment start, the Vivo 3 performs a function test indicated by a short beep and LED signals. If the function test fails, take the Vivo 3 out of use and contact your supplier of the Vivo 3

Indications of a successful function test:

- A short beep indicating successful alarm signalling
- The alarm LED first lights yellow, then red
- The audio pause LED lights yellow
- In about a second, both LEDs are turned off


6.3 Stop the Treatment

It is possible to choose how the machine will start and stop. If you want to confirm the starting and stopping or not, see3.5.5.1 *Device Settings*, page 29

Confirm Start/Stop is set to On

1 Press and hold the **On/Off** button until the *Stopping Treatment* progress bar is filled.


⇒A request for a complementary action is now displayed for stopping the treatment. If no complementary action is taken within 6 seconds, the Vivo 3 reverts to normal operation.





2 Do one of the following:

- Press the **Mute Alarm** button (a). (Always available.)
- Press **Ramp** (b) for a ramped stop. (Available if a *Ramp Down* time has been specified.)



Confirm Start/Stop is set to Off

1 Press the **On/Off** button.

2 Press **Ramp** (b) for a ramped stop. (Available if a *Ramp Down* time has been specified.)


6.4 Switch Off the Vivo 3

Confirm Start/Stop is set to On

1 Make sure that the treatment is stopped and the Vivo 3 is in *Standby* mode.


2 Press the **On/Off** button.

⇒When the message "Do you want to turn off the ventilator?" is displayed, confirm by pressing the **Mute Alarm** button within 6 seconds. Otherwise, the Vivo 3 will revert to standby mode.



3 Press the **Mute Alarm** button.

⇒ The Vivo 3 is now turned off.



Confirm Start/Stop is set to Off

1 Make sure that the Vivo 3 is in *Standby* mode.

2 Press and hold the **On/Off** button.

⇒ The Vivo 3 is now turned off.

6.5 Monitor Treatment

1 Press the **Monitor** navigation button.

⇒The Monitor page is now displayed. See 3.5.2 *The Monitor Page*, page 27

6.5.1 Treatment Values Monitored by the Vivo 3

- **Flow**

  Flow is the constant flow for HFNT. It is the only monitored value when HFNT is active.

- **Ppeak**

  P_{peak} *(Peak pressure)* is the highest pressure that is recorded during the latest inspiratory phase.

- **EPAP**

  EPAP (Expiratory Positive Airway Pressure) is the lowest pressure that is recorded during the latest expiratory phase.

- **Leakage**

  Leakage is the average calculated leak (l/min) over the last breath, with a breath by breath update.

- **MVe**

  Mve (Minute Volume, expiratory) is calculated as the Tidal Volume multiplied with the Total Breath Rate.

- **Vte**

  Vte (Tidal volume, expiratory) is the expired tidal volume for each breath.

- **SpO2**

  SpO_2 (Saturation of Peripheral Oxygen) displays the patient's oxygen saturation. as measured by the SpO_2 module.

  This value is only displayed if the SpO_2 module is connected.

  Displayed ranges: According to the manufacturer's specifications.

- **Pulse Rate**

  Pulse Rate displays the patients pulse rate as measured by the SpO2 module.

  This value is only displayed if the SpO_2 module is connected.

  Displayed ranges: According to the manufacturer's specifications.

6.6 Using Accessories

This section describes how to use accessories provided by Breas Medical.

6.6.1 Using the Attachable Humidifier



WARNING!

Read the section 2.7 *Humidification and Heating — Warnings and Precautions*, page 17 before using the Vivo 3 with the attachable humidifier.

The humidifier is for single patient use only.

NOTE

The ventilator can also be used with active stand-alone and passive humidifiers

About the Attachable Humidifier

The humidifier is intended to humidify the patient air. It is intended for non-invasive use only.


The humidifier is intended for stationary use and requires a connection to the Mains power supply to work. During mobile use of the Vivo 3, remove the humidifier water chamber and insert the air bypass unit.


Using the Attachable Humidifier for the First Time — Overview

1 Take out the air bypass unit by pressing the locking latch (a) and then pulling it out.

2 Fill the humidifier with water.

3 Insert the humidifier.






Detach the Humidifier from the Vivo 3

1 If any treatment is running, stop it.

2 Push down the locking latch (a) and then pull the humidifier out.

3 If you will use the Vivo 3 without the humidifier, install the air bypass unit in place of the humidifier.



Fill the Humidifier

Duration of operation between humidifier refills


Humidifier level (5): 12 hours. At lower settings, the duration will be longer.

CAUTION!

- The water chamber must be detached from the Vivo 3 when filling water into the water chamber.
- Use only distilled or sterilized water or boiled, chilled tap water in the humidifier water chamber. This is to reduce bacteria and mineral deposits. Do not add any substances to the water, as this can have adverse effect.
- Do not fill the water chamber with hot water.
- Do not overfill the water chamber. The water chamber has a capacity of 350 ml and the maximum filling level is indicated on the chamber.
- After using the ventilator, wait one minute before you open the water chamber since it can get hot under certain conditions (for example if the humidifier runs out of water).

1 Detach the water chamber, see above.

2 Inspect the water chamber for damages, dirt or deposits. Clean if required, see 8.1.2 *Clean the Humidifier Water Chamber*, page 71. If the water chamber is damaged, replace it before use.



3 Make sure the water meets the quality requirements. It shall be either:

- Distilled
- Sterilized
- Boiled and chilled tap water.


4 Hold the humidifier with the air path openings (a) up and fill water into either of the air path openings. Make sure not to overfill (b).

5 If the outside of the humidifier is wet, dry it with a lint free cloth before attaching it to the Vivo 3.

Attach the Humidifier to the Vivo 3

1 If any treatment is running, stop it.

2 If the air path bypass unit is installed to the Vivo 3, remove it by pressing the locking latch (a) and then pulling it out.



3 Make sure the humidifier is correctly filled and push it into the Vivo 3 so the locking latch is engaged.

⇒A click indicates that the humidifier is correctly installed.

Activate the Humidifier

1 Press the **Setup** button twice.

⇒The second setup page is now displayed, with humidifier settings.

2 Press the **Up arrow** button until the setting **Humidifier** is selected.

3 Press the **Right arrow** button to turn the humidifier on.

⇒The value is shifted to On and the humidifier symbol is lit on the display.

⇒The setting for Humidifier Level is displayed.

4 Press the **Down arrow** to select the **Humidifier Level** setting.


5 Press the **Left arrow** button to decrease the humidification or the **Right arrow** button to increase the humidification.

When not Using the Humidifier

1 Take out the humidifier by pushing down the locking latch (a) and then pull the humidifier out.

2 Empty the humidifier of water.

3 Insert the air path bypass unit.



6.6.2 Using the Heated Circuit

WARNING!

Read section 2.7 *Humidification and Heating — Warnings and Precautions*, page 17 before using the Vivo 3 with the heated circuit.

NOTE

The heated circuit requires connection to the Mains power supply to work.

Information about safety, warnings, product description, installation, usage, cleaning, maintenance and technical specifications can be found in the user instruction for the heated circuit.

For information about connecting a patient circuit with heated circuit, see 5.4 *Connecting the Patient Circuit*, page 38.

Activate the heated circuit

1 Press the **Setup** button twice.

⇒The second setup page is now displayed, with heated circuit settings.

2 Press the **Up arrow** button until the setting **Heated Circuit** is selected.

3 Press the **Right arrow** button to turn the heated circuit on.

⇒The value is shifted to On and the heated circuit symbol is lit on the display.

⇒The setting for Heated Circuit Level is displayed.

4 Press the **Down arrow** to select the **Heated Circuit Level** setting.

5 Press the **Left arrow** button to decrease the heating or the **Right arrow** button to increase the heating.

6.6.3 Using the SpO_2 Module


The SpO_2 module (006369) enables connection to an SpO_2 sensor for measuring of functional oxygen saturation of arterial haemoglobin (SpO_2) and pulse rate. The SpO_2 measurements will be stored in the data memory which can be downloaded to a PC and viewed in the PC software.

1 Connect either of the the SpO_2 adapter cable (007079) or the accessory box (007000) to the communication port (1).

2 Connect the SpO_2 module to the adapter cable or accessory box.

3 Connect the SpO_2 sensor to the SpO_2 module.

When starting the ventilator, check that the SpO_2 symbol on the screen is lit.



Technical Data

SpO2 measurement	Range: 70 to 100% Tolerance: ±3 digits. No motion and flex sensor.
Pulse rate measurement	Range: 25 to 240 bpm. Tolerance: ±3 digits. No motion and flex sensor.
SPO_2 sensor light characteristics	Peak Wavelength (red): 660 nm Peak Wavelength (infrared): 905 nm Maximum Optical Output Power: ≤ 15 mW For more information regarding the oxygen probe's characteristics and usage, please refer to the respective probe manual.

Information about safety, warnings, product description, installation, usage, cleaning, maintenance and technical specifications can be found in the user instruction for the SpO_2 module. .

CAUTION!

- When using the Vivo 3 with the SpO2 module and sensor, the Vivo 3 displays functional oxygen saturation measured by the sensor.
- Environmental factors may influence the function or accuracy of the pulse oximeter, such as ambient light, physical movement, diagnostic testing, low perfusion, electromagnetic interference, dysfunctional haemoglobin, presence of certain dyes and inappropriate positioning of the pulse oximeter probe.
- A functional test cannot be used to assess the accuracy of a pulse oximeter probe or a pulse oximeter.

6.6.4 Using the Oxygen Supply Adapter

WARNING!

Read the section 2.9 *Oxygen Usage — Warning and Precautions*, page 19 before using the Vivo 3 with oxygen.

Supplemental oxygen can be added by an oxygen source with rotameter such as oxygen cylinder, central oxygen supply system or an oxygen concentrator.

Supplemental Oxygen Supply Requirements

Property	Requirement
Maximum flow	30 l/min.
Maximum pressure	100 kPa
Supply source	Source equipped with rotameter. Examples of supply sources: • Oxygen cylinder • Central oxygen supply • Oxygen concentrator
Connector	The oxygen source shall be equipped with the Breas Low Pressure Oxygen Adapter, art. no. 005032

Connect the Oxygen supply

1 Connect the oxygen adapter (article no. 005032) to the oxygen supply's tube.

2 Connect the oxygen adapter to the oxygen port at the back of the Vivo 3. See 3.3 *Ventilator Back* , page 23 for detailed information.

3 If using a device for monitoring the oxygen saturation, connect it according to the supplier instructions.

CAUTION!

Do not turn on the oxygen supply before the therapy has been started and turn it off directly when the therapy is discontinued.

6.6.5 Using the Nurse Call Connection

The Vivo 3 can be connected to a nurse call system through a port on the accessory box.


When connected, alarms from the Vivo 3 will be forwarded to the nurse call system.

1 Connect the Accessory box to the Communication port at the back of the Vivo 3.

2 Connect the nurse call cable to the Nurse call/ Remote alarm port on the accessory box.

⇒The Nurse call symbol is now lit on the display.

3 Trigger an alarm on the Vivo 3 and check that it activates the nurse call system. For detailed information about triggering alarms, see .



6.6.6 Using the Remote Alarm Unit

Information about safety, warnings, product description, installation, usage, cleaning, maintenance and technical specifications can be found in the user instruction for Remote Alarm.


The Remote Alarm enables care providers and clinical personnel to monitor the Vivo 3 alarms remotely. The Remote Alarm forwards alarms from the Vivo 3.

1 Connect the Accessory box to the Communication port at the back of the Vivo 3.

2 Connect the remote alarm cable to the Nurse call/ Remote alarm port on the accessory box.

3 Start the remote alarm unit.

4 Trigger an alarm on the Vivo 3 and check that it activates the remote alarm system.



6.6.7 Using the Protective Cover

The protective cover is intended for mobile use of the Vivo 3 in hospitals, institutions and home care environments. It can be used while the Vivo 3 is operating, for example mounted on a wheelchair, in a personal vehicle, or carried by hand.

The protective cover protects the Vivo 3 from environmental impact such as shock, water spill, sunlight, dust and dirt, under normal handling.

CAUTION!

The protective cover does not protect the Vivo 3 from rain or snowfall.

- The protective cover can be used together with external power supply units.
- The protective cover cannot be used together with the attachable humidifier. When using the protective cover, the air path bypass unit shall be installed in place of the attachable humidifier.

6.6.8 Using the Lightweight Mobility Bag

The lightweight Mobility Bag is intended for mobile use of the Vivo 3 in hospitals, institutions and home care environments. It can be used while the ventilator is operating, for example mounted on a wheelchair, in a personal vehicle, or carried by hand.

The lightweight Mobility Bag protects the ventilator from, water spill, sunlight, dust and dirt, under normal handling.

The lightweight mobility bag cannot be used together with the attachable humidifier. When using the lightweight mobility bag, the air path bypass unit shall be installed in place of the attachable humidifier.

It does not protect against environmental impact such as shock, from rain or snowfall.

6.6.9 Using the Y-Cable

The Y-cable is used for connecting the Vivo 3 to both mains and external DC at the same time. see 5.3 *Connecting the Vivo 3 to Power Supply*, page 37. When both power sources are available, the mains will be used.


6.6.10 Using the Vivo 3 with the Trolley

Intended Use

The intended use of the trolley system is to allow patient mobility while receiving ventilator treatment. The trolley shall only be used in indoor, hospital environment. The trolley system consists of a trolley base and a mounting bracket.


This section describes how to use the Vivo 3 and a trolley with mounting bracket.

Mount and dismount the Vivo 3 as shown in the picture:



The bottom plate is mounted to the trolley using four screws.

Before mounting the bottom plate, the attachment must be dismounted:



Be careful when handling the trolley with the ventilator mounted, in order to avoid any risk of the trolley falling. The trolley can be tipped 10° and return to vertical position, when loaded in accordance with the weight specifications below.

WARNING!

The maximum total weight of the trolley and added accessories is 37 kg (82 lbs). (Trolley base weight = 12 kg (26 lbs), maximum externally added load = 25 kg (55 lbs).)

- The maximum load of the trolley basket is 0.9 kg (2 lbs).
- The maximum load of the trolley rail is 9 kg (20 lbs).

No maintenance is required.

6.7 Basic Troubleshooting

Problem	Action
The ventilator doesn't start.	Check the connection of the power cord between the ventilator and the power outlet.
The ventilator starts, but the patient circuit doesn't deliver any air.	Ensure that both ends of the patient circuit are correctly connected. Straighten the patient circuit or replace it.
The humidifier doesn't work properly.	If the humidifier is incorrectly assembled, disassemble it and then assemble it correctly. If the air is dry despite using the humidifier, increase the level of humidification. If the humidifier doesn't deliver any heat, check that the Vivo 3 is powered from the AC power supply; the humidifier doesn't work on battery.

7 Alarms

WARNING!

Risk of Unnoticed Critical Conditions

- All the physiological alarms of the Vivo 3 must be set at safe levels that will effectively warn the user of any risk.

  The alarm levels should be assessed considering the patient's treatment settings.
- Any change of treatment settings or change of components in the ventilation system may require readjustment of the alarm levels.
- The alarm sound level should be set to a clearly audible level. Setting the alarm sound level below that of the ambient sound level can impede recognition of alarm conditions.

Distributed Alarm System

Distributed alarm system for the ventilator:

- The cable connected remote alarm unit provided by Breas Medical

Alarm Function During Power Failure

At power failure, the following will happen:

1. When the mains power is broken, the internal battery takes over the power supply and the following message is displayed: Switched to Internal Battery.
2. When 20 minutes (±5 minutes) of the battery charge remains, the following alarm (Medium) is displayed: Last Power Source Low Capacity. An audio signal is heard.
3. When 5 – 7 minutes of the battery charge remains, the following alarm (High) is displayed: Critical Low Last Power Source. An audio signal is heard.
4. When 0 minutes of the battery charge remains, the following message is displayed: The internal battery is completely discharged. Connect mains or DC to charge.
5. An alarm is triggered, and the Vivo 3 is powered off.

The alarm settings are maintained during power failure.

7.1 Operator's Position

The alarm priority indications are designed to be recognized from a distance of 4 meters and by an angle of 50 ° from the normal of the Vivo 3 display.

7.1.1 Checking the Operator's Position

The patient shall not be connected during these checks.

1 With treatment started on the Vivo 3, disconnect the patient circuit to activate an alarm.

2 From the operator's position, make sure that the audible alarm signal is heard and that it is possible to recognize the alarm priority level by either of the visual signals (flashing LED lights and display message on the screen). The sound pressure level range is 55-80 dBA for high and medium priority alarm signals.

For detailed information, see 7.2.1 *Identify an Alarm Condition*, page 54.

3 If the test fails, consider the following actions:

- Find a better position for the operator.
- Adjust the alarm sound level, see 7.2.5 *Adjust the Alarm Sound Level*, page 57.

7.2 Handle Alarms

7.2.1 Identify an Alarm Condition

If an alarm condition is detected, the Vivo 3 main unit and the remote alarm unit (if connected) will alarm without alarm signal generation delay. The alarms will remain active until the alarm condition is resolved.

Active alarms are indicated by:

- Audible signal, see page 54.
- Alarm message on the screen, see page 55.
- The alarm LED, see page 55.

Alarm Audio signal

- **High priority alarms**
  3 short signals followed by 2 more after 0.5 s.
  The signal sequence repeats every 3rd second.



- **Medium priority alarms**
  3 signals. The signal sequence repeats every 6th second.



For information about adjusting the alarm sound level, see 7.2.5 *Adjust the Alarm Sound Level*, page 57.

Information Message Audio Signal

Short signal every 5 seconds



Alarm Sound Pressure Level

The alarm sound pressure level is adjustable within 55–80 dBA.

Alarm Message on the Screen

The name of the active alarm is displayed on the screen.

- **High priority alarms**
  Red highlight color.
- **Medium priority alarms**
  Yellow highlight color.

For detailed information about specific alarms, see 7.3 *Physiological Alarms*, page 57 and 7.4 *Technical Alarms*, page 65.

Display of Multiple Alarms


If several alarms are active, active high priority alarms have precedence over medium priority alarms: All high priority alarm conditions must be resolved before any medium alarms are displayed.

If several alarms of the same priority are active at the same time, the alarm descriptions are looped in the display.

A ">>" symbol indicates that more alarms are to be displayed in the loop.

Alarm LED signal


- **High priority alarms**
  Red light, flashing quickly (0.5 s. interval).
- **Medium priority alarms**
  Yellow light, flashing slowly (2 s. interval).



7.2.2 Pause the Alarm Sound

The audible signal can be paused for 60 seconds by pressing the Audio Pause button. Pressing the button again reactivates the sound.

If a new alarm condition occurs during the audio pause period, the audible signal will be reactivated.



7.2.3 Reset an Alarm

To reset an alarm, correct the cause of the alarm condition.

⇒Once the cause is corrected, the alarm disappears from the display.

WARNING!

If an alarm condition cannot be corrected, take the Vivo 3 out of use and contact your supplier of the Vivo 3.

7.2.4 View Historical Alarms


To view historical alarms, press the Alarm button until Alarm/Event history is shown.



7.2.5 Adjust the Alarm Sound Level


Alarm Sound Level is found in the Alarm Settings.

1 Press the Alarm navigation button until the Alarm settings page with Alarm Sound Level is shown.







2 Press the Down or Up arrow button to select the Alarm Sound Level setting.



3 Press the Left or Right arrow buttons to adjust the sound level.



4 Press the Up arrow button to finish the adjustment by deselecting the setting.



5 When finished with the sound level adjustment, check that the alarm can be received at the operator's position, see 7.1 *Operator's Position*, page 53.

7.3 Physiological Alarms

For complete alarm information including setting ranges, see the Clinician's manual.

7.3.1 High Pressure Alarm

Property	Description
Alarm text	**High Pressure**
Priority	High
Alarm condition	A High Pressure alarm will be given when the patient pressure reaches the set High Pressure alarm limit for three consecutive breaths.
Possible cause	• Mismatch between pressure setting and alarm setting. • Coughing during inspiration. • Changes in airway resistance and or compliance.
Reset criteria	A full breath is performed with maximum pressure below the alarm limit.
Ventilator action	The Vivo 3 will continue treatment according to the current settings. The actual breath is however terminated if the High Pressure alarm limit is reached.
Setting Display	The High Pressure alarm setting is displayed by a red line in the pressure bar graph. 

7.3.2 Low Pressure Alarm

Property	Description
Alarm text	**Low Pressure**
Priority	High
Alarm condition	A Low Pressure alarm will be given when the Vivo 3 pressure fails to reach the low pressure alarm limit for 15 seconds.
Possible cause	• Disconnection of patient circuit. • Mismatch between pressure setting and alarm setting. • Leakage from the mask or other components of the patient circuit.
Reset criteria	The pressure rises above the alarm limit.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.
Setting display	The Low Pressure alarm setting is displayed by a red line in the pressure bar graph. 

7.3.3 High Breath Rate Alarm

Property	Description
Alarm text	**High Breath Rate**
Priority	Medium
Alarm condition	A High Breath Rate alarm will be given when the alarm limit has been exceeded for 15 seconds.
Possible cause	• Mismatch between the Breath Rate setting and the alarm setting. • Increased Breath Rate. • Too sensitive setting of the inspiratory trigger setting.
Reset criteria	The breath rate goes below the alarm limit.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.3.4 Low Breath Rate Alarm

Property	Description
Alarm text	**Low Breath Rate**
Priority	High
Alarm condition	A Low Breath Rate alarm will be given when the delivered total breath rate is below the alarm limit for 15 seconds.
Possible cause	• Mismatch between the Breath Rate setting and the alarm setting. • The patient cannot trigger breaths because the inspiratory trigger setting is too high. • Decrease in the patient's spontaneous breathing. • Circuit disconnection.
Reset criteria	The breath rate goes above the alarm limit.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.3.5 High Minute Volume Alarm

Property	Description
Alarm text	**High MV**
Priority	Medium
Alarm condition	A High Minute Volume alarm will be given when the minute volume is above the set alarm limit for 15 seconds.
Possible cause	• Disconnection of patient circuit. • Mismatch between pressure setting and alarm setting. • Leakage from the mask or other components of the patient circuit. • Increased breath rate.
Reset criteria	The minute volume goes below the alarm limit.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.3.6 Low Minute Volume Alarm

Property	Description
Alarm text	**Low MV**
Priority	High
Alarm condition	A Low Minute Volume alarm will be given when the minute volume is below the set alarm limit for 15 1 seconds.
Possible cause	• Disconnection of patient circuit. • Mismatch between pressure setting and alarm setting. • Leakage from the mask or other components of the patient circuit. • Decreased breath rate.
Reset criteria	The minute volume goes above the alarm limit.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.3.7 Rebreathing Alarm

Property	Description
Alarm text	**Rebreathing**
Priority	High
Alarm condition	A Rebreathing alarm will be given if the intentional leakage is too low for more than 15 seconds.
Possible cause	• Obstructed or occluded patient circuit. • Incorrect patient circuit.
Reset criteria	The leakage is back within limits.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.3.8 Apnea Alarm

The Vivo 3 is not intended for use as an Apnea monitor.

Property	Description
Alarm text	**Apnea**
Priority	High
Alarm condition	An Apnea alarm will be given when no patient-triggered breath is detected for the set period of time.
Possible cause	• Patient stopped breathing. • Patient decreases spontaneous breathing. • Circuit disconnection. • Inspiratory Trigger is set too high.
Reset criteria	Inspiratory effort detected by the Vivo 3.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.3.9 Disconnection Alarm

CAUTION!

⚠

No single alarm can reliably detect all disconnections due to the number of possible combinations of therapy settings, circuit configurations and patient interfaces. To verify that patient disconnection can be detected, including if the patient interface becomes accidentally detached from the patient, it is advised to test the functionality of the Disconnection Alarm upfront with the complete set-up as used during treatment, including items such as filters, circuit, connectors, interface (mask, cannula etc.)

Property	Description
Alarm text	**Disconnection**
Priority	High
Alarm condition	A Disconnection alarm will be given when the measured flow exceeds the expected leakage flow at the set Pressure for 15 seconds.
Possible cause	• Too high leakage in the patient circuit. • The patient has removed the mask. • Circuit disconnection.
Reset criteria	The leakage is back within limits.
Ventilator action	The Vivo 3 will continue treatment according to the current settings

7.3.10 High EPAP Alarm

Property	Description
Alarm text	**High EPAP**
Priority	Medium
Alarm condition	A High EPAP alarm will be given when EPAP has gone 30% above the set value for 3 breaths.
Possible cause	• Blocked leakage port. • Too short expiratory time. • Changes in airway resistance and or compliance.
Reset criteria	EPAP has gone below the alarm limit (lower than 30% above the set value).
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.3.11 Low EPAP Alarm

Property	Description
Alarm text	**Low EPAP**
Priority	Medium
Alarm condition	A Low EPAP alarm will be given when EPAP has gone 30% below the set value for 3 breaths.
Possible cause	• Excessive leakage.
Reset criteria	EPAP has gone above the alarm limit (higher than 30% below the set value).
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.3.12 High SpO_2 Alarm

Property	Description
Alarm text	**High SpO2**
Priority	Medium
Alarm condition	A High SpO_2 alarm will be given when the measured SpO_2 exceeds the alarm limit for 30 seconds.
Possible cause	Too high flow of bleed-in oxygen.
Reset criteria	The SpO_2 value goes back below the alarm limit.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

This alarm requires a connected SpO_2 sensor.

7.3.13 Low SpO_2 Alarm

Property	Description
Alarm text	**Low SpO2**
Priority	High
Definition	A Low SpO_2 alarm will be given when the measured SpO_2 is below the alarm limit for 30 seconds.
Possible cause	• Too low flow of bleed-in oxygen. • Oxygen inlet is disconnected. • Delivered tidal volumes are too small.
Ventilator action	The ventilator will continue treatment with the same settings.

This alarm requires a connected SpO_2 sensor.

7.3.14 High Pulse Rate Alarm

Property	Description
Alarm text	**High Pulse Rate**
Priority	Medium
Alarm condition	A High Pulse Rate alarm will be given when the measured pulse rate exceeds the alarm limit for 15 seconds.
Possible cause	• Insufficient ventilatory support. • Too low flow of bleed-in oxygen. • The EPAP value is set too high. • Bad positioning of the finger probe.
Reset criteria	The pulse rate goes back below the alarm limit.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

This alarm requires a connected SpO_2 sensor.

7.3.15 Low Pulse Rate Alarm

Property	Description
Alarm text	**Low Pulse Rate**
Priority	High
Alarm condition	A low pulse rate alarm will be given when the measured pulse rate goes below the alarm limit for 15 seconds.
Possible cause	• Bad positioning of the finger probe. • Too low flow of bleed-in oxygen. • Insufficient ventilatory support.
Reset criteria	The pulse rate goes back above the alarm limit.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

This alarm requires a connected SpO_2 sensor.

7.3.16 Circuit Disconnection Alarm

Property	Description
Alarm text	**Circuit Disconnection**
Priority	High
Alarm condition	Only available when HFNT is active. The alarm will be given if the device detects that the patient circuit is disconnected from the device. It will not detect if the patient is connected or not, only if the cannula and/or tube is connected to the device.
Possible cause	The circuit is disconnected.
Reset criteria	The device detects that the tube and/or cannula is connected again.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.3.17 Blocked Circuit Alarm

Property	Description
Alarm text	**Blocked Circuit**
Priority	High
Alarm condition	Only available when HFNT is active. The alarm will be given if the device detects that the patient circuit is blocked and no air can be delivered.
Possible cause	The circuit is blocked.
Reset criteria	The device detects that the circuit is not blocked anymore.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.4 Technical Alarms

7.4.1 High Pressure Limitation Alarm

Property	Description
Alarm text	**High Pressure Limitation**
Priority	High
Alarm condition	The High Pressure Limitation alarm is given if the high pressure alarm limitation limit (60 cmH_2O) is reached.
Possible cause	• Mismatch between pressure setting and alarm setting. • Coughing during inspiration. • Changes in airway resistance and or compliance.
Reset criteria	A full breath with pressure below the set alarm limit.
Ventilator action	The current breath is terminated and then theVivo 3 will continue treatment according to the current settings.

7.4.2 Power Fail Alarm

Property	Description
Alarm text	The display is turned off. The Power Fail alarm is indicated by the alarm LED and the audible signal only.
Priority	High The display is turned off. The Power Fail alarm is indicated by the alarm LED and the audible signal only.
Alarm condition	The Power Fail alarm is given if the last power source fails to provide enough power for running the ventilator.
Possible cause	The last available power source cannot deliver power to the ventilator. Battery discharged or battery failure.
Reset criteria	External power supply connected to ventilator.
Ventilator action	The Vivo 3 stops the treatment, turns off the display and gives the Power Fail alarm for at least 2 minutes. If power is restored within the alarm time, the ventilator will automatically resume treatment with current settings. When powered up again, the power failure will be logged.

7.4.3 SpO_2 Sensor Failure / Disconnected Alarm

This alarm requires a connected SpO_2 sensor.

Property	Description
Alarm text	**SpO2 Disconnected**
Priority	Medium
Alarm condition	An SpO_2 Sensor Failure/Disconnection alarm will be given if one of the conditions below appears: • An error signal is received from the sensor • No signal at all from the sensor is received within 2 seconds.
Possible cause	Faulty or disconnected sensor.
Reset criteria	Normal communication with the sensor is re-established. An information message remains until acknowledged by the user.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.4.4 SpO_2 Artifact

This alarm requires a connected SpO_2 sensor.

Property	Description
Alarm text	**Poor SpO2 Signal**
Priority	Medium
Alarm condition	A poor SpO_2 Signal alarm will be given if perfusion is too low or artifacts are detected by the sensor.
Possible cause	Check the sensor and its placement on the patient.
Reset criteria	An OK signal is received from the sensor or the sensor is disconnected. An information message remains until acknowledged by the user.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.

7.4.5 Ambient Pressure Compensation Lost Alarm

Property	Description
Alarm text	**Pressure Comp Lost**
Priority	Medium
Alarm condition	An Ambient Pressure Compensation Lost alarm will be given when the automatic ambient pressure compensation functionality is out of order.
Ventilator action	The Vivo 3 will continue treatment according to the current settings. Normal atmospheric pressure at sea level will be used as approximation for the temporary ambient pressure compensation. If used at other altitude, delivered and measured pressures may deviate.
Reset	Reset of ventilator.

7.4.6 Low Battery Alarm

Property	Description
Alarm text	**Low Battery. Connect to Mains.**
Priority	Medium
Alarm condition	This alarm will be given when the last battery source (internal battery) has 15 minutes of operating time left with current settings.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.
Reset Criteria	Connection of another power source within 15 minutes or the treatment will shut down.

7.4.7 Critically Low Battery Alarm

Property	Description
Alarm text	**Critically Low Battery. Connect to Mains**
Priority	High
Alarm condition	This alarm will be given when the last battery source (internal battery) has 5 minutes of operating time left with current settings.
Ventilator action	The Vivo 3 will continue treatment according to the current settings.
Reset Criteria	Connection of another power source within 5 minutes or the treatment will shut down.

7.4.8 High Patient Air Temp. (High Patient Air Temperature)

Property	Description
Alarm text	**High Patient Air Temp**
Priority	High
Alarm condition	A High Patient Air Temperature alarm will be given when the patient air temperature exceeds 43°C (109.4°F). If having a patient circuit with an active heated circuit, the patient air temperature is measured by the circuit's temperature sensor. Otherwise, the temperature is measured by the flow measurement sensor inside the Vivo 3.
Possible cause	• Blocked air inlets. • Too high ambient temperature.
Ventilator action	The Vivo 3 will continue treatment. If a heated circuit or attachable humidifier is used, these will be turned off.
Reset criteria	The temperature goes below the limit again.

7.4.9 Flow Sensor Failure

Property	Description
Alarm text	**Flow Sensor Failure**
Priority	Medium
Alarm condition	No data or erroneous data from the flow sensor
Possible cause	
Reset criteria	Correct data from the sensor is received again. An information message remains until acknowledged by the user.
Ventilator action	The Vivo 3 will continue treatment but with the following limitations: • Monitoring of leakage is disabled. • Volume measurements are disabled. • The patient cannot trigger breaths (applies to assisted modes) An information message about the limitations is displayed on the screen.

7.4.10 Internal Function Failure

Property	Description
Alarm text	**Int. Function Failure**
Priority	High
Alarm condition	Failure of internal function that prevents treatment or normal operation of the ventilator. The error code that follows the alarm text indicates the kind of function failure.
Reset criteria	Correct function is restored.
Ventilator action	The ventilator will stop the treatment.
Action to take	Restart the Vivo 3. If the alarm persists or reoccurs: Take a note of the error code and contact your supplier of the Vivo 3.

8 Cleaning and Maintenance

This chapter contains instructions for cleaning and maintenance actions that can be carried out by the care provider or by users with physical ability and working knowledge of the system.

WARNING!

Risk of Personal Injury

- Repairs, upgrades and modifications must be carried out by technicians authorized by Breas Medical only and in accordance with instructions from Breas Medical
- The Vivo 3 must not be opened, repaired or modified by unauthorized personnel.

  If subjected to unauthorized operations, Breas Medical is no longer responsible for the performance and safety of the device and all warranties will become invalid.

The patient-connected parts and the filters must be cleaned and replaced regularly to ensure correct function of the Vivo 3. All replaced parts must be disposed of in accordance with local environmental regulations regarding the disposal of used equipment and waste.

8.1 Cleaning the Vivo 3

WARNING!

Risk of Electric Shock

Disconnect the power supply before cleaning theVivo 3 according to the instructions in this manual.

Do not soak the Vivo 3 or immerse it into any fluids.

8.1.1 Clean the Main Unit Externally

Equipment

- A lint free cloth.
- A mild soap solution or Ethanol 70%.

1 Turn off the Vivo 3 and disconnect the power supply.

2 Remove the patient circuit.

3 If any cable connected accessories (like the SpO_2 sensor or the accessory box) are used, disconnect them.

4 Clean the outside of the Vivo 3 using a lint free cloth moistened with a mild soap solution and / or ethanol 70% for surface disinfection.

5 Clean the optional accessories in accordance with their respective user instructions.

6 Allow all cleaned parts to air dry.

7 Visually inspect the ventilator and accessories that have been cleaned. If the ventilator or any of the accessories are not visually clean and dry, repeat the associated cleaning steps or safely dispose of the device.

8 Reconnect the patient circuit and any accessories that was disconnected during the cleaning.

8.1.2 Clean the Humidifier Water Chamber

The water chamber shall be cleaned weekly.

WARNING!

Risk of Burns
After using the ventilator, wait one minute before you open the water chamber since it can get hot under certain conditions (for example if the humidifier runs out of water).


1 Open the humidifier by pulling the lid up. Don't detach the silicone airpaths (a). Also, don't dismount the cartridge heater (b).

2 Wash the humidifier by hand with mild soap and warm water, or in a dishwasher without dishwashing detergent.

3 Allow all cleaned parts to air dry.

4 Visually inspect the cleaned parts. If they are not visually clean and dry, repeat the associated cleaning steps or safely dispose of the device.

5 After cleaning, ensure that the silicone gasket (c) is positioned correctly: When closing the humidifier; the grooved edge must be positioned downwards.



8.1.3 Clean the Patient Circuit

CAUTION!

The cleaning and replacement intervals should be established by the care provider, based on the care provider's infection control procedures and the instructions from the patient circuit's manufacturer.
Check the patient circuit regularly for damage. In case of damage, replace the circuit.

The patient circuit should be cleaned and replaced in accordance with the manufacturer's instructions, or by the care provider's instructions if additional cleaning or other replacement intervals are prescribed.

8.2 Clean and Replace Patient Air Inlet Filters


The Vivo 3 patient air inlet filters are located inside a magnetic filter holder at the back of the ventilator. The table below describes the filters and their minimum maintenance intervals.

Filter	Maintenance Intervals (minimum)*
Air inlet filter, grey (coarse)	• Wash: every week. • Replace: every year or when assigning the Vivo 3 to a new patient.
Air inlet filter, white (fine)	• Replace: every 4th week or when assigning the Vivo 3 to a new patient.

* If the Vivo 3 is used in an environment with high grades of pollen or pollutions, shorter intervals might be required.

1 Turn off the Vivo 3 and place it on a clean dust free surface.


2 Pull out the filter holder and remove the filters.



3 If required by the interval or if visibly dirty, wash the grey coarse filter:

1. Wash the filter using warm water and a mild soap.
2. Rinse thoroughly.
3. Dry the filter by first squeezing it in a towel and then letting it dry in the air. Do not wring the filter.
4. Check that the filter is visually clean, completely dry and undamaged before reinstalling it. If not visually clean, repeat the cleaning. If damaged, dispose of the filter safely and replace it with a new one.

4 When reinstalling the air inlet filters in the filter holder: first install the grey coarse filter, then the white fine filter.




5 Reinstall the filter holder on the Vivo 3.

8.3 Disposal

The ventilator, any accessories and all replaced parts must be disposed of and recycled in accordance with the local environmental regulations regarding the disposal of used equipment and waste. Contact your service provider for information regarding the disposal procedure.

NOTE



Batteries used with the ventilator shall be recycled in accordance with the local environmental regulations.

9 Technical Specifications

9.1 Ventilator Size and Weight

Property	Value
Dimensions (WxHxD)	With air bypass unit: 181 x 165 x 200 mm With humidifier: 181 x 165 x 240 mm
Weight	1,8 kg

9.2 Power Supply

Mains Power Supply

Property	Value
Mains Power Supply	100–240 V AC tolerance: +10%/-20%, 50 to 60 Hz, max 1.2 A. The approved AC/DC supply listed in 10 *Accessories*, page 81 must be used.
Protection against electric shock	Class II ME Equipment

External DC Power Supply

Property	Value
External DC Supply	12–24 V isolated DC The approved DC/DC supply listed in 10 *Accessories*, page 81 must be used.

Internal Battery

Property	Value
Battery type	4 cell Li-Ion battery
Nominal Voltage	14.4 V
Capacity	2.1 Ah
Operational time	3 hours
Expected life:	300 charging cycles
Charging time (Empty to full)	< 2 hours.

NOTE

When 300 cycles have been exceeded, a notification message is displayed.

9.3 Operating Conditions

Environmental Condition	Specification
Normal Operation Temperature	+5°C to +40°C (Humidifier: +15°C to +35°C) • Be careful not to position the Vivo 3 in an extra warm place, such as in direct sunlight or above a radiator. • Caution should be exercised if the room temperature is higher than 36°C (97°F). The air flow for breathing produced by the Vivo 3 can be as much as 4°C (7°F) higher than the ambient room temperature.
Extended Operation Temperature	–20°C to +5°C The Vivo 3 is operational during the extended operation temperature for 4 hours, if: • The Vivo 3 is first started within the normal operation temperature span. • The Vivo 3 is placed in it's protective cover. • This condition happens maximum once a day. • The ambient air is dry and still.
Transport and Storage Temperature	• +5°C to +45°C (Maximum 90 days) • –25°C to +70°C (Maximum 30 days) • When the ventilator is brought from minimum/maximum allowed storage temperature, ensure that it is warmed up/cooled down for one hour before starting it. • The Vivo 3 must not be stored in a warm place, such as direct sunlight or close to a radiator. If stored in temperatures outside normal operation conditions, let the Vivo 3 acclimate before taking it to use.
Humidity	RH from 15% to 95%, non-condensing.
Ambient Pressure Range	70 to 106 kPa This corresponds to ~ 315 m below sea level to ~3000 m above sea level

Environmental Condition	Specification
Ventilator ingress protection	IP 22 Mechanical ingress protection: protected from touch by fingers and objects greater than 12.5 mm Liquid ingress protection: The device withstands dripping water(equivalent to 3 mm rainfall /minute) when not tilted more than 15 degrees from vertical. The protection has been tested for 10 minutes (2.5 minutes in every tilt direction).

Mechanical ingress protection: protected from touch by fingers and objects greater than 12.5 mm
Liquid ingress protection: The device withstands dripping water(equivalent to 3 mm rainfall /minute) when not tilted more than 15 degrees from vertical. The protection has been tested for 10 minutes (2.5 minutes in every tilt direction).

Precaution

- There is a silicone lid to protect the USB, SD card, and communication ports. The IP22 classification is applicable only when this lid is in place. However, the accessory box can be connected with retained IP22 classification, but then only the lower part of the silicone lid can be opened.
- Ensure that the silicone lid on the back of the ventilator is closed when no accessories are connected.

9.4 Emission and Immunity Declaration

The essential performance listed below has been tested according to applicable parts of IEC 60601-1-2.

Vivo 3 Essential Performance

The ventilator will deliver ventilation at the patient-connection port within its published accuracy specifications and within the alarm limits set by the operator, or generate any of the following alarms:

- High pressure
- Low pressure
- Low inspired minute volume
- High inspired minute volume
- Low breath rate
- High leakage
- Low last power source
- Power failure
- Disconnection
- High EPAP
- Low EPAP

The ventilator will provide SpO2 and pulse rate values within its published accuracy specifications and generate an alarm upon a low SpO2 condition. The ventilator will provide indication when the SpO2 value or pulse rate is potentially incorrect, and generate an alarm condition to indicate when the SpO2 value update period has exceeded 30 seconds.

Under the immunity test conditions, the following allowances are acceptable:

- Temporary degradation of performance that does not adversely affect basic safety or essential performance.
- Any temporary degradation of SpO2 performance following transient immunity test exposure shall recover from any disruption within 30 seconds.

Additionally, the following shall not be allowed:

- permanent damage or unrecoverable loss of function
- changes in programmable parameters or settings
- reset to default settings
- change of operating mode
- initiation of unintended operation
- negative false alarm condition

9.4.1 Guidance and Manufacturer's Declaration – Electromagnetic Immunity

The ventilator is intended for use in the electromagnetic environment specified below. The customer or the user of the ventilator should assure that it is used in such an environment.

Immunity test	Compliance level	Electromagnetic environment — Guidance
Electrostatic discharge (ESD) IEC 61000-4-2	±8 kV contact ±15 kV air	The relative humidity should be at least 5%.
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply lines ±1 kV for input/output lines	Mains power quality should be that of a typical commercial, hospital and residential environment.
Surge IEC 61000-4-5	± 0.5, ± 1 kV, ± 2 kV line to line	Mains power quality should be that of a typical commercial, hospital and residential environment.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	30 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial, hospital and residential environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	0% U_T, 0.5 cycle (multiple phase analysis); 0% U_T, 1 cycle; 70% U_T, 25/30 cycles (50/60 Hz); 0% U_T, 250/300 cycles (50/60 Hz);	The ventilator runs on internal battery during voltage dips, short interruptions and voltage variations on power supply input lines.
Proximity Magnetic Fields IEC 61000-4-39	30 kHz: 8 A/m 134.2 kHz: 65 A/m 13.56 MHz: 7.5 A/m	Proximity magnetic fields should be at levels characteristic of a typical location in a typical commercial, hospital, and residential environment.

NOTE

U_T is the mains voltage prior to application of the test level.

WARNING!

Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the ventilator, including specified cables. Otherwise, degradation of the performance of this equipment could result.

Immunity test	IEC 60601 test level	Recommended separation distance
Conducted RF IEC 61000-4-6	3 V_{rms} (6 V_{rms} inside ISM/ASR bands)	d=0.35*√P m at 150 kHz to 80 MHz
Radiated RF IEC 61000-4-3	10 V/m 80 MHz to 2.7 GHz	d= 1.2*√P m at 80 MHz to 800 MHz d= 2.3*√P m at 800 MHz to 2.7 GHz **Equation description:** P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey[a], should be less than the compliance level in each frequency range[b]. Interference may occur in the vicinity of equipment marked with this symbol.

- At 80 MHz and 800 MHz, the higher frequency range applies.
- These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

a) Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the ventilator is used exceeds the applicable RF compliance level above, the ventilator should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the ventilator.

b) Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 10 V/m.

Proximity fields from RF wireless communications equipment

Frequency Range and Level: RF wireless communication equipment		
Test Frequency (MHz)	**Modulation***	**Immunity Level (V/m)**
385	Pulse Modulation: 18 Hz	27
450	Pulse Modulation: 18 Hz	28
710 745 780	Pulse Modulation: 217 Hz	9
810 870 930	Pulse Modulation: 18 Hz	28
1720 1845 1970	Pulse Modulation: 217 Hz	28
* The carrier shall be modulated using a 50 % duty cycle square wave signal.		

Frequency Range and Level: RF wireless communication equipment		
Test Frequency (MHz)	**Modulation***	**Immunity Level (V/m)**
2450	Pulse Modulation: 217 Hz	28
5240 5500 5785	Pulse Modulation: 217 Hz	9
* The carrier shall be modulated using a 50 % duty cycle square wave signal.		

9.4.2 Guidance and Manufacturer's Declaration – Electromagnetic Emission

The ventilator is intended for use in the electromagnetic environment specified below. The customer or the user of the ventilator should assure that it is used in such an environment.

Immunity test	Compliance	Electromagnetic environment – Guidance
RF emissions CISPR 11	Group 1	The ventilator uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	The ventilator is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations/flicker emission IEC 61000-3-3	Complies	

9.4.3 Recommended separation distances between portable and mobile RF communications equipment and the ventilator

The ventilator is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the ventilator can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the ventilator as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter (W)	Separation distance according to the frequency of transmitter (m)		
	150 kHz to 80 MHz $d = 0.35*\sqrt{P}$ m	80 MHz to 800 MHz $d = 1.2*\sqrt{P}$ m	800 MHz to 2.7 GHz $d = 2.3*\sqrt{P}$ m
0.01	0.035	0.12	0.23
0.1	0.11	0.38	0.73
1	0.35	1.20	2.3
10	1.1	3.8	7.3
100	3.5	12.0	23

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

Notes

- At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.
- These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

9.4.4 Recommended separation distances between external power conductors and the ventilator

Rated maximum current in conductor (A)	Separation distance (m)
	50-60 Hz: d= I/2πH= I/188
1	0.005
10	0.05
30	0.16

For conductors rated at a maximum current not listed above, the recommended separation distance d in meters (m) can be estimated using the equation d=I/2πH, where I is the maximum current rating of the conductor in amperes (A) according to the transmitter manufacturer; H is the ventilator immunity compliance level to electromagnetic fields in the 50-60 Hz frequency span (30 A/m).

10 Accessories

CAUTION!


Responsibility for System

Accessory equipment connected to the analogue and digital interfaces must be certified according to the respective IEC standards (e.g. IEC 60950 or IEC 62368 for data processing equipment and IEC 60601-1 for medical equipment). Furthermore all configurations must comply with the valid version of the system standard IEC 60601-1. Anybody who connects additional equipment to the signal input part or signal output part is configuring a medical system, and is therefore responsible for ensuring that the system complies with the requirements of the valid version of the system standard IEC 60601-1. If in doubt, consult the technical service department or your local representative.
A PC, when connected to the ventilator, shall comply with IEC 62368-1, IEC 60950-1 or IEC 60601-1.

10.1 Patient Circuits and Air Delivery Accessories


Circuit: 22 mm smoothbore with leak port

Function: Delivers air to the patient, applied part
Part No: 005060




Circuit: 1.8m x 22mm Smoothbore disposable

Function: Delivers air to the patient, applied part
Part No: 009118



Circuit: Single limb 22 mm. Single patient multiple use.


Function: Delivers air to the patient, applied part
Part No: 008426 (30-pack of 004465)



Circuit: 1.8m x 15mm Smoothbore Disposable

Function: Deliver air to the patient


Part No: 009119



Circuit: Single limb 15 mm. Single patient multiple use.

Function: Delivers air to the patient, applied part


Part No: 008427 (30–pack of 006712)



Heated Circuit

Function: Delivers air to the patient, applied part. Prevents rain-out.

Part No: 006990




Leakage Port

Function: Providing a leakage for clearing exhaled gases.

Part No: 007243 (10 pieces)

Low resistance bacterial filter, with CO_2 connector



Function: Filter air at ventilator outlet

Part No: 007963

Characteristics

- BFE (Bacterial Filtration Efficiency): 99.9935%
- VFE (Viral Filtration Efficiency): 99.9608 %
  Deadspace: 36 ml
- Tidal volume range: 150–1.500 ml

Resistance (cmH_2O pressure drop)

	30 l/min	60 l/min	90 l/min
Insp.	1.0	2.3	4.0
Exp	0.9	1.9	3.3

Oxygen Adapter, Low Pressure

Function: Connection of oxygen supply.

Part No: 005032

10.2 Power Accessories

DC/DC Power Supply (RRC CAR 70M)

Function: Power supply adapter for the ventilator. Shall be used when connecting to an external DC source. Compliant with IEC 60601–1.


Part No: 006995

Y Cable

Function: Power supply cable, for connecting to both AC and DC power supply.

Part No: 007006

XPAC - External battery with charger



Function: Extends usage time of supported Breas products.


Part No Cable for connection to device: 007803

Part No Charger with cable:
Single: Charger with one battery
Dual: Charger with two batteries

Single: 007995, Dual: 007999

10.3 Ventilator Filters and Detachable Parts

Attachable humidifier




Function: Humidifies the patient air. For non-invasive use only.

Part No: 006977

Air Bypass Unit

Function: Directs the patient air flow, if the attachable humidifier is not used.


Part No: 006983



Filter Holder

Function: Holder for air inlet filters


Part No: 007598



Air Inlet Filter, Grey

Function: Coarse air inlet filter, user replaceable part. Long life (washable).


Part No: 007203 (5 pieces)



Air Inlet Filter, White

Function: Fine air inlet filter, user replaceable part. Disposable.

Part No: 007202 (5 pieces)




10.4 Monitoring Accessories

Accessory Box

Function: For connecting measurement and communication accessories:

- Nurse call cable or Remote alarm
- SpO_2 sensor (Might also be connected directly to the Vivo 3, if no other measurement or communication accessories are used.)


Part No: 007000



Remote alarm with cable

Function: Monitor Vivo 3 alarms remotely

Part No: 10 m: 006348, 25 m: 006349



Nurse call cable

Function: Connect the ventilator to a hospital nurse call system


Part No:

NO: 006365

NC: 006364

10 kΩ, NO: 006363

10 kΩ, NC: 006362



SpO_2 module

Function: Connection interface

Part No: 006369




SpO_2 sensor

Function: Finger Clip SpO_2 sensor

Part No:

Adult: 006589


Paediatric: 006590



SpO_2 sensor

Function: Multisite SpO_2 sensor

Part No: 006591



SpO_2 Adapter Cable

Function: Connection cable. For use of SpO_2 without accessory box.

Part No: 007079



PC Software

Function: Support software for follow-up on patient treatment.

Part No: 007067

USB Cable

Function: USB cable for transferring data between a PC and the Vivo 3.

Part No: 005757

Memory card

Function: Storage and transfer of settings, patient data and usage data


Part No: 006705



Memory card reader/writer

Function: Read/write memory card

Part No: 002185



10.5 Other Accessories

Protective Cover

Function: Shock protection

Part No: 007014


Lightweight Mobility Bag

Function: For mobile use of the Vivo 3 in hospital, institutions and home care environments.

Part No: 007380


Breas Trolley

Function: Mobile use, transportation
Part No: 007384




Mounting Bracket

Function: Bracket to mount the Vivo 3 to the trolley
Part No: 006998



Universal rail clamp

Function: Attach a humidifier to a trolley.
Part No: 007858



Accessory Bag

Function: Storage of accessories
Part No: 007989



Appendices

A FAA Compliance

To whom it may concern:

In line with the FAA Advisory Circular AC 91.21-1D October 27, 2017, this kind of respiratory assistive device may be used onboard an aircraft, without further testing by the carrier, provided they have been tested for Electromagnetic Compatibility (EMC) in accordance with the current version of RTCA/DO-160, Section 21, Category M.

Breas Medical has successfully completed testing for the ventilator System. The ventilator System complies with RTCA/DO-160, Section 21, Category M and can be considered FAA compliant.

Some airlines may require advance notification before travel, and devices may need to be operated by battery. Breas Medical recommends that customers check with their airline.

FAA Compliance (English text)

To whom it may concern:

In line with the FAA Advisory Circular AC 91.21-1D October 27, 2017, this kind of respiratory assistive device may be used onboard an aircraft, without further testing by the carrier, provided they have been tested for Electromagnetic Compatibility (EMC) in accordance with the current version of RTCA/DO-160, Section 21, Category M.

Breas Medical has successfully completed testing for the ventilator System. The ventilator System complies with RTCA/DO-160, Section 21, Category M and can be considered FAA compliant.

Some airlines may require advance notification before travel, and devices may need to be operated by battery. Breas Medical recommends that customers check with their airline.

Index

A

AC Voltage
supply 73
Accessories 81
using 44
AHI 31
Air outlet 23
Alarm 30
Ambient Pressure Compensation Lost 67
Apnea 61
Circuit 65
Critically Low Battery. Connect to Mains 68
Disconnection 62, 65
Flow sensor failure 69
High breath rate 59
High EPAP 62
High minute volume 60
High Patient Air Temp (High Patient Air Temperature) 68
High pressure 58
High pressure limitation 65
High pulse rate 64
High SpO_2 63
Internal Function Failure 69
Low Battery 67
Low breath rate 59
Low EPAP 63
Low minute volume 60
Low pressure 58
Low pulse rate 64
Low SpO_2 63
Power Failure Alarm 66
Rebreathing 61
resetting 55
SpO_2 artifact 67
SpO_2 sensor failure / disconnected 66
Alarm condition
identifying 54
Alarm message 55
Alarm signal
Check 39
LED 55
Alarm sound
pausing 55
Alarm time 30
Alarm volume 30
Alarms 53
physiological 57
Ambient Pressure Compensation Lost alarm 67
Apnea alarm 61
Average usage hours 31

B

Battery 23
Battery percentage 30
Battery time left 30
Battery,(internal)
specifications 73
Blocked Circuit 65
Board revision 30
Bootloader version 30
Buttons 22

C

Caution
Icon 8
Check
Compliance and resistance 39
Start-up 39
Circuit Disconnection 65
Cleaning 18, 70
Clock 30
Command
execute 26
Communciation port 23
Compliance
Check 39
Contraindications 7
Critically Low Battery. Connect to Mains alarm 68

D

Date 30
Date format 30
Days with usage 31
DC Voltage
supply 73
Device information 30
Device memory 30
Device settings 29
Disconnection alarm 62

Display 22
Display light 29

E

Electromagnetic compatibility 13
Electromagnetic emission 79
Electromagnetic immunity 77
Electrostatic discharge 13
Environmental conditions 74
Erase memory data 30

F

Filter
  usage 15
Filters
  clean, replace 71
Firmware version 30
Front panel 22

H

Heated circuit
  setting 36
  using 47
Heating 17
High breath rate alarm 59
High EPAP alarm 62
High Patient Air Temperature Alarm (High Patient Air Temp) alarm 68
High pressure alarm 58
High pulse rate alarm 64
High SpO_2 alarm 63
Humidification 17
Humidifier
  setting 36
Humidifier, attachable
  using 44
Hypoventilation 59

I

Icon
  Caution 8
  Note 8
  Reference 8
  Warning 8
Interface board revision 30
Internal Function Failure alarm 69
Invasive use 16

L

Languages 29
Leakage
  Check 39
Light intensity 29
Low Battery alarm 67
Low breath rate alarm 59
Low EPAP alarm 63
Low pressure alarm 58
Low pulse rate 64
Low SpO_2 alarm 63

M

Maintenance 18, 70
Memory card port 23
Memory data 30
Menus 24
Min daily use 31
Mobile use
  safety 20
Monitor treatment 44

N

Note
  Icon 8
Nurse call connection
  using 50

O

Operating conditions 74
Oxygen 49
  usage 19
Oxygen concentration 20
Oxygen port 23

P

Patient air outlet 23
Patient circuit
  connect 38
Placing 37
Power Failure Alarm alarm 66
Power handling
  Check 39
Power management 23

Power port 23
Power supply
connect 37
Pre-use test 39
Pre-Use Test
Failure 40
Pressure unit 29
Product name 30
Protective cover 50

R

Ramp down 36
Ramp up 35
Rebreathing alarm 61
Reference
Icon 8
Remote alarm unit
using 50
Reset compliance data 31
Resistance
Check 39
RF communications equipment 79

S

Save memory data on card 30
SD card port 23
Settings 35
change 26
Setup pages 28
Sound level, alarm 57
SpO_2 sensor
using 48
Start date 31
Stop the treatment 42
Switch off 43
Switch on 41
Symbols 31

T

Technical specifications 73
Time 30
Time format 30
Total days 31
Total usage hours 31
Transit operation
safety 20
Treatment
start 41
Treatment data 27

U

Undesirable side effects 7
Usage time
Internal battery 73
USB port 23
Using instructions 41

V

Voltage
supply 73

W

Warning
Icon 8
Wheel chair
safety 20

Y

Y-cable 51